From: DEP-Overflowmicro
To: "June Flowers"
Cc: phil@flowerslabs.com; Savage, Jacqueline
Subject: FW: Your FCL report is ready.
Date: Monday, July 10, 2017 1:42:00 PM

Ms. June,

Could you please let me know if all the wells fluoresced for this sample? I am not sure If the result needs to be "J" qualified.

Thank You,

Carina Tull
Biological Scientist I
FDEP Laboratory - Biology Program
Florida Department of Environmental Protection
Carina.Tull@dep.state.fl.us
Office: 850.245.8174

-----Original Message-----
From: labinfo@flowerslabs.com [mailto:labinfo@flowerslabs.com]
Sent: Wednesday, June 28, 2017 1:38 PM
To: Matthews, Daisys <Daisys.Matthews@dep.state.fl.us>; Ayala, Natalie <Natalie.Ayala@dep.state.fl.us>; DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Your FCL report is ready.

June 28, 2017

Thank you for doing business with Flowers Chemical Laboratories! Results for your samples sampled on June 27, 2017 with Project Number SMP is available at http://www.flowerslabs.com/Apps/WebObjects/FCLSite.woa/wa/arpt?ar=kWWshytcKzhgzD8

Please note that Adobe Acrobat Reader is required to view this report. If you don't already have this program, it may be downloaded at http://www.adobe.com/products/acrobat/readstep2.html.

From: June Flowers
To: DEP-Overflowmicro
Cc: Savage, Jacqueline; Cyndi LaRosa
Subject: 335923SW1 - no "J"
Date: Friday, July 14, 2017 4:40:54 PM

Carina,

The "J" flag was added in error. We reported from the 100mL, because the 10mL was much lower.
49 large and 47 small well fluoresced.

We will remove the "J" and revise the report on Monday.
Thank you,

June Flowers
Quality Systems Director
Flowers **C**hemical **L**aboratories, Inc.
481 Newburyport Ave., Altamonte Springs FL 32701
PO Box 150597, Altamonte Springs FL 32715-0597
407.339.5984 x212
www.flowerslabs.com

Celebrating 60 years 1957 - 2017

On Jul 13, 2017, at 5:07 PM, DEP-Overflowmicro wrote:

> Ms. June,
>
> Can you please add the following comment to the "J" qualified sample? "All Quanti-Tray wells were positive. Result may be higher than reported."
>
> Thank you!
>
> Carina Tull
> Biological Scientist I
> FDEP Laboratory - Biology Program
> Florida Department of Environmental Protection
> Carina.Tull@dep.state.fl.us
> Office: 850.245.8174
>
> -----Original Message-----
> From: labinfo@flowerslabs.com [mailto:labinfo@flowerslabs.com]
> Sent: Thursday, June 29, 2017 3:24 PM
> To: Matthews, Daisys <Daisys.Matthews@dep.state.fl.us>; DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
> Subject: Your FCL report is ready.
>
> June 29, 2017

Thank you for doing business with Flowers Chemical Laboratories! Results for your samples sampled on June 28, 2017 with Project Number SMP G5 Pick-up is available at
http://www.flowerslabs.com/Apps/WebObjects/FCLSite.woa/wa/arpt?ar=5z9FxNDDACrTq

Please note that Adobe Acrobat Reader is required to view this report. If you don't already have this program, it may be downloaded at
http://www.adobe.com/products/acrobat/readstep2.html.
[Dep Customer Survey]<http://survey.dep.state.fl.us/?refemail=DEP-Overflowmicro@dep.state.fl.us>