

Chemical Analysis Report

CEN-DIST-2017-07-28-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Spring Garden Crk / Deep Crk**
Request ID: **RQ-2017-07-24-77**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 16-AUG-2017 16:50



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Spring Garden Creek @ just DS of Deep Creek Collection Date/Time: 07/27/2017 10:58

Field ID: G2CE0126

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1917422	EPA 180.1 Rev. 2.0	Turbidity	3.4		NTU	P329346	
	EPA 300.0 Rev. 2.1	Chloride	190		mg Cl/L	P329389	
		Sulfate	39		mg SO4/L	P329388	
	SM 2120 B	Color (true)	160		PCU	P329349	
	SM 2320 B-97	Total Alkalinity	103		mg CaCO3/L	P329527	
	SM 2540 C-97	TDS	584		mg/L	P329771	
	SM 2540 D-97	TSS	5	I	mg/L	P329862	
	SM 4500 F-C-97	Fluoride	0.096	I	mg F/L	P329530	
1917424	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P329570	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P329446	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P329440	
	EPA 365.1 Rev. 2.0	Total-P	0.062		mg P/L	P329463	
	SM 5310 B-00	Organic Carbon	16		mg C/L	P329792	
1917426	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	I	mg P/L	P329356	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Deep Creek @ 170m US of Spring Garden Creek

Collection Date/Time: 07/27/2017 11:04

Field ID: G2CE0083

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1917423	EPA 180.1 Rev. 2.0	Turbidity	1.6		NTU	P329346	
	EPA 300.0 Rev. 2.1	Chloride	23		mg Cl/L	P329389	
		Sulfate	0.93		mg SO4/L	P329388	
	SM 2120 B	Color (true)	1.3E+03		PCU	P329349	
	SM 2320 B-97	Total Alkalinity	3.3		mg CaCO3/L	P329527	
	SM 2540 C-97	TDS	221		mg/L	P329771	
	SM 2540 D-97	TSS	2	U	mg/L	P329862	
	SM 4500 F-C-97	Fluoride	0.058	I	mg F/L	P329530	
1917425	EPA 350.1 Rev. 2.0	Ammonia-N	0.062		mg N/L	P329570	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.7		mg N/L	P329446	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.011		mg N/L	P329440	
	EPA 365.1 Rev. 2.0	Total-P	0.058		mg P/L	P329463	
	SM 5310 B-00	Organic Carbon	93		mg C/L	P329792	
1917427	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.032		mg P/L	P329356	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P329346

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329388

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329570

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329446

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329440

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329463

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329356

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P329349

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P329527

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P329771

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P329862

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P329530

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P329792

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329388

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	92.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	93.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329570

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329446

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P329446

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	106		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P329440

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P329463

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P329356

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
 Batch ID: P329527

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
 Batch ID: P329530

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.4		P	92 - 103

Reference Method: SM 5310 B-00
 Batch ID: P329792

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P329388

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917284	Sulfate	94.5		P	90 - 110
1917345	Sulfate	94.4		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917284	Chloride	96.1		P	90 - 110
1917345	Chloride	96.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329570

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917415	Ammonia-N	92.2	99.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329440

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917332	NO2NO3-N	99.0	98.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329463

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917477	Total-P	97.1	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329356

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917356	O-Phosphate-P	97.6	97.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P329530

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917482	Fluoride	99.5	96.3	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P329792

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917416	Organic Carbon	95.5	94.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P329346

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917410	Turbidity	0.0	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329388

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917284	Sulfate	0.813	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917284	Chloride	0.577	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329570

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917415	Ammonia-N	5.86	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329446

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917414	Kjeldahl Nitrogen	1.78	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329440

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917332	NO2NO3-N	0.506	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329463

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917477	Total-P	3.76	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329356

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917356	O-Phosphate-P	0.205	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P329349

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917348	Color (true)	0.784	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P329527

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917482	Total Alkalinity	0.0694	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P329771

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917411	TDS	4.92	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P329530

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917482	Fluoride	2.44	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P329792

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917416	Organic Carbon	0.565	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A77948
Included Lab Sample IDs: 1917422, 1917423

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	103	P/P	90 - 110
Sulfate	105	102	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A78030
Included Lab Sample IDs: 1917422, 1917423

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.0	P/P	95 - 105

Reference Method: SM 2120 B
Run ID: A78031
Included Lab Sample IDs: 1917422, 1917423

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	103	P/P	85 - 115

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A78036

Included Lab Sample IDs: 1917426, 1917427

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.7	98.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A78075

Included Lab Sample IDs: 1917424, 1917425

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A78099

Included Lab Sample IDs: 1917424

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.5	99.2	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A78103

Included Lab Sample IDs: 1917422, 1917423

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A78103

Included Lab Sample IDs: 1917422, 1917423

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	96.5	98.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A78118

Included Lab Sample IDs: 1917425

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.8	99.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A78129

Included Lab Sample IDs: 1917424, 1917425

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	101	P/P	90 - 110
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A78134

Included Lab Sample IDs: 1917424, 1917425

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.4	95.9	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00

Run ID: A78233

Included Lab Sample IDs: 1917424, 1917425

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.3	97.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.0	
EPA 300.0 Rev. 2.1	Chloride	93.7	96.1 96.1		0.577	
	Sulfate	92.4	94.5 94.4		0.813	
EPA 350.1 Rev. 2.0	Ammonia-N	102	92.2 99.0			5.86
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	106				1.78
EPA 353.2 Rev. 2.0	NO2NO3-N	104	99.0 98.5			0.506
EPA 365.1 Rev. 2.0	Total-P	98.0	97.1 101			3.76
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101	97.6 97.8			0.205
SM 2120 B	Color (true)				0.784	
SM 2320 B-97	Total Alkalinity	103			0.0694	
SM 2540 C-97	TDS				4.92	
SM 4500 F-C-97	Fluoride	95.4	99.5 96.3			2.44
SM 5310 B-00	Organic Carbon	96.0	95.5 94.4			0.565

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1917422, 1917423
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1917422, 1917423
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1917422, 1917423
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1917424, 1917425
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1917424, 1917425
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1917424, 1917425
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1917424, 1917425
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1917426, 1917427
SM 2120 B / E31780	True Color measured at 450 nm	1917422, 1917423
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1917422, 1917423
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1917422, 1917423
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1917422, 1917423
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1917422, 1917423
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1917424, 1917425

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/28/2017			07/28/2017 15:25	DeAsia Armster	1917422, 1917423
EPA 300.0 Rev. 2.1	07/28/2017			07/28/2017	Julia Storbeck	1917422, 1917423
EPA 350.1 Rev. 2.0	07/28/2017			08/02/2017	Sofia Elnemr	1917424, 1917425
EPA 351.2 Rev. 2.0	07/28/2017	08/01/2017	Adrian Shelby	08/02/2017	Joseph Suarez	1917424, 1917425
EPA 353.2 Rev. 2.0	07/28/2017			08/01/2017	Beverly Y. Sanders	1917424, 1917425
EPA 365.1 Rev. 2.0	07/28/2017	08/01/2017	Sima Feizi	08/02/2017	Lipi Saha	1917424, 1917425
EPA 365.1 Rev. 2.0 dissolved	07/28/2017			07/28/2017 16:00	Ping Hua	1917426
	07/28/2017			07/28/2017 16:01	Ping Hua	1917427
SM 2120 B	07/28/2017			07/28/2017 15:35	Tanya Welborn	1917422, 1917423
SM 2320 B-97	07/28/2017			08/02/2017	Dale L. Simmons	1917422, 1917423
SM 2540 C-97	07/28/2017			08/01/2017	Yijie Li	1917422, 1917423
SM 2540 D-97	07/28/2017			08/01/2017	Yijie Li	1917422, 1917423
SM 4500 F-C-97	07/28/2017			08/02/2017	Dale L. Simmons	1917422, 1917423
SM 5310 B-00	07/28/2017			08/04/2017	Sofia Elnemr	1917424, 1917425