

Chemical Analysis Report

CEN-DIST-2017-08-09-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **LVI (SMP)**
Request ID: **RQ-2017-08-07-34**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 23-AUG-2017 11:07



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Weir @ Center

Collection Date/Time: 08/08/2017 11:32

Field ID: G1CE0064

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1919895	EPA 180.1 Rev. 2.0	Turbidity	2.2		NTU	P329943	
	EPA 300.0 Rev. 2.1	Chloride	53		mg Cl/L	P329985	
		Sulfate	16		mg SO4/L	P329987	
	SM 2120 B	Color (true)	4.7	I	PCU	P329952	
	SM 2320 B-97	Total Alkalinity	21		mg CaCO3/L	P330131	
	SM 2540 C-97	TDS	128		mg/L	P330169	
	SM 2540 D-97	TSS	4	I	mg/L	P330571	
	SM 4500 F-C-97	Fluoride	0.21		mg F/L	P330133	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P330139	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.71		mg N/L	P330074	
1919896	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P330100	
	EPA 365.1 Rev. 2.0	Total-P	0.012		mg P/L	P330037	
	SM 5310 B-00	Organic Carbon	6.5		mg C/L	P330165	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	U	mg P/L	P329961	
	dissolved						

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P329943

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P329985

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P329987

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P330139

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P330074

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P330100

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P330037

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329961

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P329952

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P330131

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P330169

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P330571

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P330133

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P330165

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329985

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329987

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P330139

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P330074

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	107		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P330100

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	106		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P330037

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329961

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.1		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P330131

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P330133

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.4		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P330165

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.3		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329985

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1919548	Chloride	99.0		P	90 - 110
1919711	Chloride	96.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329987

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1919548	Sulfate	98.1		P	90 - 110
1919711	Sulfate	95.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P330139

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1919799	Ammonia-N	102	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P330100

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1919542	NO2NO3-N	100	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P330037

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1919937	Total-P	92.4	94.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329961

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1919897	O-Phosphate-P	92.0	93.3	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P330133

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1920503	Fluoride	97.4	97.9	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P330165

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1919562	Organic Carbon	99.3	99.5	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P329943

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919891	Turbidity	0.803	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329985

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919711	Chloride	6.48	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329987

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919711	Sulfate	1.53	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P330139

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919799	Ammonia-N	1.47	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P330074

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919901	Kjeldahl Nitrogen	5.03	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P330100

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919542	NO2NO3-N	0.623	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P330037

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919937	Total-P	1.90	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329961

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919897	O-Phosphate-P	1.40	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P329952

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919793	Color (true)	0.757	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P330131

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1920503	Total Alkalinity	0.611	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P330169

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919825	TDS	3.98	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P330133

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1920503	Fluoride	0.492	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P330165

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1919562	Organic Carbon	0.194	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A78270

Included Lab Sample IDs: 1919895

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	106	105	P/P	90 - 110
Sulfate	102	104	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A78298

Included Lab Sample IDs: 1919895

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: SM 2120 B

Run ID: A78299

Included Lab Sample IDs: 1919895

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.6	100	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A78303

Included Lab Sample IDs: 1919897

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.2	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A78348

Included Lab Sample IDs: 1919896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.5	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A78349

Included Lab Sample IDs: 1919896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	101	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A78364

Included Lab Sample IDs: 1919895

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.3	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A78364

Included Lab Sample IDs: 1919895

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	105	98.1	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A78366

Included Lab Sample IDs: 1919896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A78375

Included Lab Sample IDs: 1919896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A78402

Included Lab Sample IDs: 1919896

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	99.2	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity					0.803	
EPA 300.0 Rev. 2.1	Chloride	100	96.3	99.0		6.48	
	Sulfate	99.2	95.5	98.1		1.53	
EPA 350.1 Rev. 2.0	Ammonia-N	101	102	101			1.47
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	107					5.03
EPA 353.2 Rev. 2.0	NO2NO3-N	106	100	102			0.623
EPA 365.1 Rev. 2.0	Total-P	97.8	92.4	94.2			1.90
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.1	92.0	93.3			1.40
SM 2120 B	Color (true)					0.757	
SM 2320 B-97	Total Alkalinity	102				0.611	
SM 2540 C-97	TDS					3.98	
SM 4500 F-C-97	Fluoride	96.4	97.4	97.9			0.492
SM 5310 B-00	Organic Carbon	98.3	99.3	99.5			0.194

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1919895
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1919895
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1919895
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1919896

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1919896
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1919896
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1919896
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1919897
SM 2120 B / E31780	True Color measured at 450 nm	1919895
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1919895
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1919895
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1919895
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1919895
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1919896

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/09/2017			08/09/2017 15:43	DeAsia Armster	1919895
EPA 300.0 Rev. 2.1	08/09/2017			08/09/2017	Julia Storbeck	1919895
EPA 350.1 Rev. 2.0	08/09/2017			08/10/2017	Sofia Elnemr	1919896
EPA 351.2 Rev. 2.0	08/09/2017	08/11/2017	Adrian Shelby	08/15/2017	Joseph Suarez	1919896
EPA 353.2 Rev. 2.0	08/09/2017			08/11/2017	Beverly Y. Sanders	1919896
EPA 365.1 Rev. 2.0	08/09/2017	08/10/2017	Sima Feizi	08/11/2017	Lipi Saha	1919896
EPA 365.1 Rev. 2.0 dissolved	08/09/2017			08/09/2017 14:01	Ping Hua	1919897
SM 2120 B	08/09/2017			08/09/2017 15:28	Tanya Welborn	1919895
SM 2320 B-97	08/09/2017			08/12/2017	Dale L. Simmons	1919895
SM 2540 C-97	08/09/2017			08/10/2017	Yijie Li	1919895
SM 2540 D-97	08/09/2017			08/10/2017	Yijie Li	1919895
SM 4500 F-C-97	08/09/2017			08/12/2017	Dale L. Simmons	1919895
SM 5310 B-00	08/09/2017			08/12/2017	Sofia Elnemr	1919896