

Chemical Analysis Report

CEN-DIST-2017-08-29-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Lk Irma + LVI**
Request ID: **RQ-2017-08-28-25**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 15-SEP-2017 15:21

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Irma @ Center

Collection Date/Time: 08/28/2017 11:15

Field ID: G2CE0104

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1925909	EPA 180.1 Rev. 2.0	Turbidity	1.1		NTU	P331023	
	EPA 300.0 Rev. 2.1	Chloride	19		mg Cl/L	P331238	
		Sulfate	6.0		mg SO4/L	P331232	
	SM 2120 B	Color (true)	28	A	PCU	P331020	
	SM 2320 B-97	Total Alkalinity	20		mg CaCO3/L	P331108	
	SM 2540 C-97	TDS	85		mg/L	P331276	
	SM 2540 D-97	TSS	3	I	mg/L	P331268	
	SM 4500 F-C-97	Fluoride	0.070	I	mg F/L	P331109	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	I	mg N/L	P331165	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.56		mg N/L	P331033	
1925910	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P331113	
	EPA 365.1 Rev. 2.0	Total-P	0.024		mg P/L	P331055	
	SM 5310 B-00	Organic Carbon	7.7		mg C/L	P331190	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.004	U	mg P/L	P331003	
	dissolved						

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P331023

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P331232

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P331238

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P331165

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P331033

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P331113

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P331055

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P331003

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P331020

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P331108

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P331276

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P331268

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P331109

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P331190

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P331232

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P331238

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P331165

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	100		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P331033

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	96.7		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P331113

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P331055

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.7		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P331003

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.7		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P331108

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P331109

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.1		P	92 - 103

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P331190

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	103		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P331232

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1925903	Sulfate	100		P	90 - 110
1926043	Sulfate	95.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P331238

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1925903	Chloride	106		P	90 - 110
1926043	Chloride	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P331165

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1925910	Ammonia-N	96.3	94.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P331033

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1925910	Kjeldahl Nitrogen	104	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P331113

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1925929	NO2NO3-N	104	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P331055

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1925910	Total-P	98.2	96.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P331003

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1925911	O-Phosphate-P	95.8	97.9	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P331109

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1924584	Fluoride	96.7	97.9	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P331190

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1924475	Organic Carbon	103	102	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P331023

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925926	Turbidity	4.32	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P331232

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925903	Sulfate	4.84	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P331238

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925903	Chloride	5.88	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P331165

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925910	Ammonia-N	1.76	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P331033

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925910	Kjeldahl Nitrogen	0.813	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P331113

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P331113

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925929	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P331055

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925910	Total-P	1.28	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P331003

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925911	O-Phosphate-P	2.17	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P331020

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925909	Color (true)	1.54	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P331108

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1924584	Total Alkalinity	0.457	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P331276

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1925884	TDS	3.23	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P331109

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1924584	Fluoride	0.996	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P331190

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1924475	Organic Carbon	0.860	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A78683

Included Lab Sample IDs: 1925911

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.7	98.7	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A78689

Included Lab Sample IDs: 1925909

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.0	101	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A78690

Included Lab Sample IDs: 1925909

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	100	P/P	95 - 105

Reference Method: SM 2320 B-97

Run ID: A78722

Included Lab Sample IDs: 1925909

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	99.3	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A78722

Included Lab Sample IDs: 1925909

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	97.0	97.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A78723

Included Lab Sample IDs: 1925910

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	99.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A78728

Included Lab Sample IDs: 1925910

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.8	106	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A78733

Included Lab Sample IDs: 1925910

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A78741

Included Lab Sample IDs: 1925910

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	98.0	94.8	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A78746

Included Lab Sample IDs: 1925910

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	108	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A78767

Included Lab Sample IDs: 1925909

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	108	P/P	90 - 110
Sulfate	101	109	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
 Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
 Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				4.32	
EPA 300.0 Rev. 2.1	Chloride	100	106	102	5.88	
	Sulfate	96.6	100	95.7	4.84	
EPA 350.1 Rev. 2.0	Ammonia-N	100	96.3	94.5		1.76
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	96.7	104	103		0.813
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	104	104		0.0
EPA 365.1 Rev. 2.0	Total-P	97.7	98.2	96.9		1.28
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.7	95.8	97.9		2.17
SM 2120 B	Color (true)				1.54	
SM 2320 B-97	Total Alkalinity	102			0.457	
SM 2540 C-97	TDS				3.23	
SM 4500 F-C-97	Fluoride	95.1	96.7	97.9		0.996
SM 5310 B-00	Organic Carbon	103	103	102		0.860

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1925909
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1925909
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1925909
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1925910
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1925910
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1925910
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1925910
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1925911
SM 2120 B / E31780	True Color measured at 450 nm	1925909
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1925909
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1925909
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1925909
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1925909
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1925910

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/29/2017			08/29/2017 16:39	DeAsia Armster	1925909
EPA 300.0 Rev. 2.1	08/29/2017			09/01/2017	Julia Storbeck	1925909
EPA 350.1 Rev. 2.0	08/29/2017			08/31/2017	Sofia Elnemr	1925910
EPA 351.2 Rev. 2.0	08/29/2017	08/30/2017	Adrian Shelby	08/31/2017	Joseph Suarez	1925910
EPA 353.2 Rev. 2.0	08/29/2017			08/31/2017	Beverly Y. Sanders	1925910
EPA 365.1 Rev. 2.0	08/29/2017	08/30/2017	Sima Feizi	08/31/2017	Lipi Saha	1925910
EPA 365.1 Rev. 2.0 dissolved	08/29/2017			08/29/2017 13:52	Megan Treadwell	1925911
SM 2120 B	08/29/2017			08/29/2017 15:35	Tanya Welborn	1925909
SM 2320 B-97	08/29/2017			08/30/2017	Dale L. Simmons	1925909
SM 2540 C-97	08/29/2017			08/30/2017	Yijie Li	1925909
SM 2540 D-97	08/29/2017			08/30/2017	Yijie Li	1925909
SM 4500 F-C-97	08/29/2017			08/30/2017	Dale L. Simmons	1925909
SM 5310 B-00	08/29/2017			09/01/2017	Ping Hua	1925910