

Chemical Analysis Report

CEN-DIST-2017-09-27-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Cow Creek**
Request ID: **RQ-2017-09-25-57**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 13-OCT-2017 16:30

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Cow Creek at Osteen-Maytown Road

Collection Date/Time: 09/26/2017 09:56

Field ID: G2CE0086

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1931804	EPA 180.1 Rev. 2.0	Turbidity	2.1		NTU	P332342	
	EPA 300.0 Rev. 2.1	Chloride	19		mg Cl/L	P332363	
		Sulfate	0.20	U	mg SO4/L	P332366	
	SM 2120 B	Color (true)	1.3E+03		PCU	P332371	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P332632	
	SM 2540 C-97	TDS	189		mg/L	P332921	
	SM 2540 D-97	TSS	3	I	mg/L	P332914	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P332636	RPD
1931806	EPA 350.1 Rev. 2.0	Ammonia-N	0.17		mg N/L	P332821	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.6		mg N/L	P332504	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.008	U	mg N/L	P332611	
	EPA 365.1 Rev. 2.0	Total-P	0.063		mg P/L	P332460	
	SM 5310 B-00	Organic Carbon	85		mg C/L	P332516	
1931808	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.013		mg P/L	P332329	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 353.2 Rev. 2.0: NO2NO3-N: The MDL was elevated due to sample matrix interference.

Sample Location: Silver Lake @ Center

Collection Date/Time: 09/26/2017 11:18

Field ID: G2CE0122

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1931805	EPA 180.1 Rev. 2.0	Turbidity	0.90		NTU	P332342	
	EPA 300.0 Rev. 2.1	Chloride	45		mg Cl/L	P332364	
		Sulfate	23		mg SO4/L	P332367	
	SM 2120 B	Color (true)	12		PCU	P332371	
	SM 2320 B-97	Total Alkalinity	8.0	A	mg CaCO3/L	P332632	
	SM 2540 C-97	TDS	120	A	mg/L	P332921	
	SM 2540 D-97	TSS	2	IA	mg/L	P332914	
	SM 4500 F-C-97	Fluoride	0.044	IJ	mg F/L	P332636	RPD
1931807	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P332758	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.41		mg N/L	P332504	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P332611	
	EPA 365.1 Rev. 2.0	Total-P	0.011		mg P/L	P332460	
	SM 5310 B-00	Organic Carbon	4.8		mg C/L	P332516	
1931809	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.006	I	mg P/L	P332329	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-97: Refer to the narrative for an explanation of QC Codes.

Quality Assurance Report

Method Blank Results

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P332342

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332363

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332364

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332366

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332367

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332758

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332821

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332504

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332611

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332460

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332329

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P332371

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P332632

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P332921

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P332914

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P332636

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P332516

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332363

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332363

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	96.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332364

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332366

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332367

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332758

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332821

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332504

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.0		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332611

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	107		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332460

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.9		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332329

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	104		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P332632

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P332636

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.2		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P332516

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	101		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332363

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931610	Chloride	93.3		P	90 - 110
1931643	Chloride	94.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332364

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931910	Chloride	105		P	90 - 110
1932028	Chloride	99.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332366

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931610	Sulfate	94.7		P	90 - 110
1931643	Sulfate	93.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332367

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931910	Sulfate	104		P	90 - 110
1932028	Sulfate	101		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332758

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931807	Ammonia-N	103	99.9	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332821

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1932446	Ammonia-N	101	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332504

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931818	Kjeldahl Nitrogen	98.6	95.3	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332611

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931755	NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332460

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931818	Total-P	99.3	97.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332329

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931880	O-Phosphate-P	96.1	97.3	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P332636

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931805	Fluoride	96.0	100	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P332516

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931655	Organic Carbon	96.8	102	P/P	85 - 115

Quality Assurance Report Precision

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P332342

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931742	Turbidity	1.29	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332363

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931610	Chloride	2.38	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332364

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931910	Chloride	1.87	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332366

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931610	Sulfate	3.55	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332367

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931910	Sulfate	7.60	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332758

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931807	Ammonia-N	3.05	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332821

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1932446	Ammonia-N	0.819	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332504

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931818	Kjeldahl Nitrogen	2.26	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332611

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931755	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332460

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931818	Total-P	1.37	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332329

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931880	O-Phosphate-P	0.790	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P332371

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931769	Color (true)	4.14	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P332632

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931805	Total Alkalinity	0.258	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P332921

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931805	TDS	5.00	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P332636

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931805	Fluoride	4.04	Spike	F	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P332516

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931655	Organic Carbon	3.24	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A79194

Included Lab Sample IDs: 1931808, 1931809

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	101	96.5	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A79197

Included Lab Sample IDs: 1931804, 1931805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A79203

Included Lab Sample IDs: 1931804, 1931805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	97.3	94.3	P/P	90 - 110
Chloride	99.1	97.3	P/P	90 - 110
Sulfate	101	95.8	P/P	90 - 110
Sulfate	95.8	93.8	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A79205

Included Lab Sample IDs: 1931804, 1931805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	103	103	P/P	90 - 115
Color (true)	104	102	P/P	90 - 115

Reference Method: SM 5310 B-00

Run ID: A79261

Included Lab Sample IDs: 1931806, 1931807

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.9	98.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A79263

Included Lab Sample IDs: 1931806

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A79280

Included Lab Sample IDs: 1931807

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.2	98.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A79298

Included Lab Sample IDs: 1931806, 1931807

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	104	104	P/P	90 - 110
NO2NO3-N	105	104	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A79305

Included Lab Sample IDs: 1931804, 1931805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	98.8	P/P	90 - 110
Total Alkalinity	98.8	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A79305

Included Lab Sample IDs: 1931804, 1931805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	99.5	P/P	90 - 110
Fluoride	99.5	99.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A79313

Included Lab Sample IDs: 1931806, 1931807

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	96.7	98.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A79344

Included Lab Sample IDs: 1931807

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	104	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A79368

Included Lab Sample IDs: 1931806

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				1.29	
EPA 300.0 Rev. 2.1	Chloride	96.6	93.3	94.2	2.38	
	Chloride	102	105	99.0	1.87	
	Sulfate	98.9	94.7	93.7	3.55	
	Sulfate	103	104	101	7.60	
	Ammonia-N	103	103	99.9		3.05
	Ammonia-N	102	101	100		0.819
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	98.0	98.6	95.3		2.26
EPA 353.2 Rev. 2.0	NO2NO3-N	107	104			0.0
EPA 365.1 Rev. 2.0	Total-P	98.9	99.3	97.9		1.37
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	104	96.1	97.3		0.790
SM 2120 B	Color (true)				4.14	
SM 2320 B-97	Total Alkalinity	102			0.258	
SM 2540 C-97	TDS				5.00	
SM 4500 F-C-97	Fluoride	98.2	96.0	100		4.04
SM 5310 B-00	Organic Carbon	101	96.8	102		3.24

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1931804, 1931805
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1931804, 1931805
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1931804, 1931805
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1931806, 1931807
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1931806, 1931807
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1931806, 1931807
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1931806, 1931807
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1931808, 1931809
SM 2120 B / E31780	True Color measured at 450 nm	1931804, 1931805
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1931804, 1931805
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1931804, 1931805
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1931804, 1931805
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1931804, 1931805
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1931806, 1931807

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/27/2017			09/27/2017 17:33	Tanya Welborn	1931804, 1931805
EPA 300.0 Rev. 2.1	09/27/2017			09/28/2017	Julia Storbeck	1931804, 1931805
EPA 350.1 Rev. 2.0	09/27/2017			10/04/2017	Sofia Elnemr	1931807
	09/27/2017			10/05/2017	Sofia Elnemr	1931806
EPA 351.2 Rev. 2.0	09/27/2017	09/29/2017	Adrian Shelby	10/02/2017	Tanya Welborn	1931806, 1931807
EPA 353.2 Rev. 2.0	09/27/2017			10/03/2017	Beverly Y. Sanders	1931806, 1931807
EPA 365.1 Rev. 2.0	09/27/2017	09/29/2017	Sima Feizi	10/02/2017	Lipi Saha	1931806, 1931807
EPA 365.1 Rev. 2.0 dissolved	09/27/2017			09/27/2017 15:16	Megan Treadwell	1931808
	09/27/2017			09/27/2017 15:18	Megan Treadwell	1931809
SM 2120 B	09/27/2017			09/27/2017 15:42	DeAsia Armster	1931805
	09/27/2017			09/27/2017 16:19	DeAsia Armster	1931804
SM 2320 B-97	09/27/2017			10/03/2017	Dale L. Simmons	1931804, 1931805
SM 2540 C-97	09/27/2017			09/29/2017	Yijie Li	1931804, 1931805
SM 2540 D-97	09/27/2017			09/29/2017	Yijie Li	1931804, 1931805
SM 4500 F-C-97	09/27/2017			10/03/2017	Dale L. Simmons	1931804, 1931805
SM 5310 B-00	09/27/2017			09/29/2017	Ping Hua	1931806, 1931807