

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page

Meter ID:

B10

Project:

2017-04-17-07

Wilderness Waterway

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

4/18/17

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Chase Conley	4/17/17	0658	23.2	8.5	8.6	100			P F	L F
ICV	Chase Conley	4/17/17	0701	22.9	8.6	8.6	100			P F	L F
CCV	Chase Conley	4/17/17	1635	24.2	8.4	8.2	98			P F	L F
CCV										P F	L F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Chase Conley	4/17/17	0705	170412	4/18	1000	1000	P F	L F
ICV	Chase Conley	4/17/17	0706	1604188	4/17	500	500	P F	L F
CCV	Chase Conley	4/17/17	1637	2511C38	11/17	5000	5020	P F	L F
CCV								P F	L F

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Chase Conley	4/17/17	0707	167801	12/18	7.0	7.0	-26	P F	L F
Calibr.	Chase Conley	4/17/17	0710	167431	11/18	4.0	4.0	153	P F	L F
Calibr.	Chase Conley	4/17/17	0711	165882	9/18	10.0	10.0	-201	P F	L F
ICV	Chase Conley	4/17/17	0712	167431	11/18	4.0	4.0	155	P F	L F
CCV	Chase Conley	4/17/17	1639	165882	9/18	10.00	10.00	-200	P F	L F
CCV									P F	L F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

Notes: