

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name

Meter ID:

B10

Project:

2017-05-28-03
SMP Marco

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

4/18/17

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Chase Conley	5/23/17	0720	24.1	8.4	8.4	100	53		☒ F	☒ F
ICV	Chase Conley	5/23/17	0722	24.0	8.4	8.4	100	53		☒ F	☒ F
CCV	Ben Pender	6/11	1700	26.4	8.0	8.1	100	53		☒ F	☒ F
CCV										P F	L F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Chase Conley	5/23/17	0724	170412	4/18	1000	1000	☒ F	☒ F
ICV	Chase Conley	5/23/17	0726	170412A	4/18	500	504	☒ F	☒ F
CCV	BP	1/11	1705	2511C38	11/17	50,000	50538	☒ F	☒ F
CCV								P F	L F

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Chase Conley	5/23/17	0728	167801	12/18	7.0	7.0	-33	☒ F	☒ F
Calibr.	Chase Conley	5/23/17	0730	167431	11/18	4.0	4.0	146	☒ F	☒ F
Calibr.	Chase Conley	5/23/17	0732	165882	9/18	10.0	10.0	-212	☒ F	☒ F
ICV	Chase Conley	5/23/17	0734	167431	11/18	4.0	4.0	144	☒ F	☒ F
CCV	BP	1/11	1710			10.0	9.09	-205	☒ F	☒ F
CCV									P F	L F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

Notes: