

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page

Meter ID:

18

Project:

SMP - LEE STREAMS

RQ-2017-07-24-06

Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).

(2) Always wait for meter to stabilize before recording any readings.

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 8/7/2017

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	MARYANN DUGAN	8-7-2017	0733	24	8.42	8.63	100. -	53		(P) F	(L) F
ICV	"	8-7-2017	0739	24	8.42	8.18	97	50.2		(P) F	(L) F
CCV	"	8-7-2017	1435	28.7	7.73	7.45	96.4	50.1		(P) F	(L) F
CCV										P F	L F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	MARYANN DUGAN	8-7-2017	0743	170421A	5/18	10,000	9589	(P) F	(L) F
ICV	"	8-7-2017	0745	170412A	4/18	500	514	(P) F	(L) F
CCV	"	8-7-2017	1438	170412	4/18	100	105	(P) F	(L) F
CCV								P F	L F

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	MARYANN DUGAN	8-7-2017	0748	167801	12/18	7.0	7.0	-11.8	(P) F	(L) F
Calibr.	"	8-7-2017	0752	171629	4/19	4.0	4.0	75.6	(P) F	(L) F
Calibr.	"	8-7-2017	0759	170903	2/19	10.0	10.0	-893.6	(P) F	(L) F
ICV	"	8-7-2017	0804	170903	2/19	10.0	10.0	-95.2	(P) F	(L) F
CCV	"	8-7-2017	1724	171629	4/19	4.0	3.9	77.8	(P) F	(L) F
CCV									P F	L F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

Notes: