

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID:

17

Project:

Peace River / Bobcat, Lindt

RQ-2017-08-14-61

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

8/7/17

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Chase Conley	8/17/17	1001	22.4	8.4	8.4	100	12.7		P F	L F
ICV	Chase Conley	8/17/17	1003	22.6	8.6	8.6	100			P F	L F
CCV	MARYANN DUGAN	08/17/2017	1520	29.3	7.651	7.0	100			P F	L F
CCV										P F	L F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Chase Conley	8/17/17	1005	170728A	8/18	1000	1000	P F	L F
ICV	Chase Conley	8/17/17	1008	170728B	8/18	500	502	P F	L F
CCV	MARYANN DUGAN	08/17/2017	1520	170421B	5/18	100		P F	L F
CCV	Kirk, W. 4c	8/18	905	170421B	5/18	100	84	P F	L F

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Chase Conley	8/17/17	1010	168705	2/19	7.0	7.0	-2.7	P F	L F
Calibr.	Chase Conley	8/17/17	1012	171629	4/19	4.0	4.0	169	P F	L F
Calibr.	Chase Conley	8/17/17	1014	172902	4/19	10.0	10.0	-181	P F	L F
ICV	Chase Conley	8/17/17	1016	172902	4/19	10.0	9.9	-180	P F	L F
CCV	MARYANN DUGAN	08/17/2017	1525	171629	4/19	4.0	4.0	165.9	P F	L F
CCV									P F	L F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

Notes: