

RQ-2017-07-17-06

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name

Meter ID:

18

Project:

SMP - Sarasota Stream

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

8/21/17

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Parker	8/22/17	736	24.5		8.34	100.0	42.1		P F	L F
ICV	Ben Parker	8/22/17	732	23.9		8.5	100.4	47.1		P F	L F
CCV	MARYANN DUGAN	8/22/17	1532	25.2		8.20	99.6	-65		P F	L F
CCV										P F	L F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.**Rapid-Pulse Sensors:** DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.**Optical:** DO gain range 0.85 to 1.15; DO charge N/A.**Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Parker	8/22/17	733	170728B	8/18	500	500	P F	L F
ICV	Ben Parker	8/22/17	734	170728A	8/18	1,000	996	P F	L F
CCV	MARYANN DUGAN	8/22/17	1534	170421A	5/18	10,000	9783	P F	L F
CCV			1533	170421B	5/18	100	96	P F	L F

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Parker	8/22/17	736	168205	2/19	7.0	7.0	-37	P F	L F
Calibr.	Ben Parker	8/22/17	737	171629	4/19	4.0	4.0	60.6	P F	L F
Calibr.	Ben Parker	8/22/17	738	172902	4/19	10.0	9.97	-123	P F	L F
ICV	BP	8/22/17	740	172902	4/19	10.0	9.94	-123	P F	L F
CCV	MD	8/22/17	1549	171629	4/19	4.0	4.1	57	P F	L F
CCV									P F	L F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50;
 mV pH 10 Range -180 $\pm$ 50; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater

Notes: _____