

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

keys

Project:

2017-12-11-93

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 12/04/17

DO DEP SOP FT 1500	Name	Date	Time (ET)	Temp. (°C)	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Gary Smardk	12/11/17	0708	24.93	8.3	8.3	100	52.3	0.9539	P F	L F
ICV	Gary Smardk	12/11/17	0712	24.6	8.3	8.4	100	52.3		P F	L F
CCV	JOHN BARRINGTON	12/11/17	1025	26.0	8.1	7.9	98	53.3		P F	L F
CCV										P F	L F

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time (ET)	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Gary Smardk	12/11/17	0720	170728B	8/18	500	500	P F	L F
ICV	Gary Smardk	12/11/17	0722	17082B	9/18	100	100.02	P F	L F
CCV	JOHN BARRINGTON	12/11/17	1020	170728B	8/18	500	497	P F	L F
CCV								P F	L F

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time (ET)	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Gary Smardk	12/11/17	0732	173426	5/19	7.0	6.99	-14.6	P F	L F
Calibr.	Gary Smardk	12/11/17	0735	172900	4/14	4.0			P F	L F
Calibr.	Gary Smardk	12/11/17	0735	172902	4/14	10.0	10.0	-191.0	P F	L F
ICV	Gary Smardk	12/11/17	0740	173426	5/19	7.0	6.9	-11.7	P F	L F
CCV	JOHN BARRINGTON	12/11/17	1023	173426	5/19	7.0	7.0	-16.9	P F	L F
CCV									P F	L F

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or $\pm 0.05m$, whichever is greater

Notes: