

Chemical Analysis Report

WQAP-2017-03-16-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **SMP: Clam Bay & Gulf of Mexico**
Request ID: **RQ-2017-03-13-51**
Customer: **WQAP**
Project ID: **SMP**

Send Reports to:
DEP SROC
2295 Victoria Ave.
Suite 364
Fort Myers, FL 33901
Attn: Ben Paswater

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, PhD, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 12-APR-2017 10:19

A handwritten signature in black ink, appearing to read "Colin Wright", is written over a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: 8063-2 OFFSHORE OF GORDON RIVER

Collection Date/Time: 03/15/2017 09:45

Field ID: G1SD031517-1

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1880475	EPA 180.1 Rev. 2.0	Turbidity	23		NTU	P321691	
	SM 2320 B-97	Total Alkalinity	128		mg CaCO3/L	P322239	
	SM 2540 D-97	TSS	28		mg/L	P322158	
	SM 4500 F-C-97	Fluoride	1.2		mg F/L	P322244	
1880477	EPA 350.1 Rev. 2.0	Ammonia-N	0.021		mg N/L	P321774	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.53		mg N/L	P321850	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.008	I	mg N/L	P322039	
	EPA 365.1 Rev. 2.0	Total-P	0.071		mg P/L	P321854	
	SM 5310 B-00	Organic Carbon	2.7		mg C/L	P321723	
1880479	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.010		mg P/L	P322467	

Sample Location: MIDDLE OF MIDDLE CLAM BAY

Collection Date/Time: 03/15/2017 11:00

Field ID: G1SD031517-2

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1880476	EPA 180.1 Rev. 2.0	Turbidity	11		NTU	P321691	
	SM 2320 B-97	Total Alkalinity	135		mg CaCO3/L	P322239	
	SM 2540 D-97	TSS	17	A	mg/L	P322158	
	SM 4500 F-C-97	Fluoride	1.2		mg F/L	P322244	
1880478	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P321774	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.34	I	mg N/L	P321850	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P322039	
	EPA 365.1 Rev. 2.0	Total-P	0.065		mg P/L	P321854	
	SM 5310 B-00	Organic Carbon	3.6		mg C/L	P321723	
1880480	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.019		mg P/L	P322467	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P321691

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P321774

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P321850

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322039

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P321854

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322467

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-97
Batch ID: P322239

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 D-97
Batch ID: P322158

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P322244

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P321723

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P321774

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P321850

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P321850

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322039

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P321854

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	99.1		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322467

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.2		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P322239

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P322244

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.3		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P321723

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P321774

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1880112	Ammonia-N	103	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P321850

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P321850

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1880477	Kjeldahl Nitrogen	103	98.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322039

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1880112	NO2NO3-N	102	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P321854

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1879507	Total-P	106	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322467

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1880527	O-Phosphate-P	96.0	96.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P322244

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1881572	Fluoride	95.9	96.5	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P321723

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1880477	Organic Carbon	96.2	87.1	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P321691

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1880466	Turbidity	10.9	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P321774

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1880112	Ammonia-N	1.04	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P321850

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1880477	Kjeldahl Nitrogen	4.09	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P322039

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1880112	NO2NO3-N	1.47	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P321854

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1879507	Total-P	0.793	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P322467

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1880527	O-Phosphate-P	0.780	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P322239

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1881572	Total Alkalinity	0.0349	Sample	P	0 - 4.8

Reference Method: SM 2540 D-97
Batch ID: P322158

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1880476	TSS	2.35	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P322244

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1881572	Fluoride	0.473	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P321723

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1880477	Organic Carbon	9.25	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A75149
Included Lab Sample IDs: 1880475, 1880476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	100	P/P	95 - 105

Reference Method: SM 5310 B-00
Run ID: A75160
Included Lab Sample IDs: 1880477, 1880478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.4	97.2	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A75184
Included Lab Sample IDs: 1880477, 1880478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A75239
Included Lab Sample IDs: 1880477, 1880478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	98.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A75289
Included Lab Sample IDs: 1880477, 1880478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	101	P/P	90 - 110
Kjeldahl Nitrogen	92.9	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A75295
Included Lab Sample IDs: 1880477, 1880478

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	101	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A75378
Included Lab Sample IDs: 1880475, 1880476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	104	106	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97

Run ID: A75378

Included Lab Sample IDs: 1880475, 1880476

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	96.7	96.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A75403

Included Lab Sample IDs: 1880479, 1880480

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.3	P/P	90 - 110
O-Phosphate-P	100	99.6	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				10.9	
EPA 350.1 Rev. 2.0	Ammonia-N	96.0	103 101			1.04
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	103 98.0			4.09
EPA 353.2 Rev. 2.0	NO2NO3-N	102	102 103			1.47
EPA 365.1 Rev. 2.0	Total-P	99.1	106 105			0.793
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.2	96.0 96.8			0.780
SM 2320 B-97	Total Alkalinity	104			0.0349	
SM 2540 D-97	TSS				2.35	
SM 4500 F-C-97	Fluoride	95.3	95.9 96.5			0.473
SM 5310 B-00	Organic Carbon	99.6	96.2 87.1			9.25

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1880475, 1880476
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1880477, 1880478
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1880477, 1880478
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1880477, 1880478
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1880477, 1880478
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1880479, 1880480
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1880475, 1880476
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1880475, 1880476
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1880475, 1880476
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1880477, 1880478

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	03/16/2017			03/16/2017 14:13	Sima Feizi	1880475, 1880476
EPA 350.1 Rev. 2.0	03/16/2017			03/17/2017	Sofia Elnemr	1880477, 1880478
EPA 351.2 Rev. 2.0	03/16/2017	03/21/2017	Adrian Shelby	03/23/2017	Joseph Suarez	1880477, 1880478
EPA 353.2 Rev. 2.0	03/16/2017			03/23/2017	Beverly Y. Sanders	1880477, 1880478
EPA 365.1 Rev. 2.0	03/16/2017	03/21/2017	Adrian Shelby	03/21/2017	Lipi Saha	1880477, 1880478
EPA 365.1 Rev. 2.0 dissolved	03/16/2017			03/16/2017 14:15	Ping Hua	1880479
	03/16/2017			03/16/2017 14:42	Ping Hua	1880480
SM 2320 B-97	03/16/2017			03/24/2017	Dale L. Simmons	1880475, 1880476
SM 2540 D-97	03/16/2017			03/20/2017	Sima Feizi	1880475, 1880476
SM 4500 F-C-97	03/16/2017			03/24/2017	Dale L. Simmons	1880475, 1880476
SM 5310 B-00	03/16/2017			03/17/2017	Julie Michelle Frank	1880477, 1880478