

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site:

SMP RA-2017-01-23-32

Date:

1/26/17

Meter #

DEP 132237

Temperature (Quarterly)

For Date of Last Temperature Verification see

in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL ICV CCV		MM	1/26	0733			8.76	22.3	101	8.93	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard µmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading µmhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL ICV CCV		MM	1/26	0740	1413	10/4/17	KC1 K-8			1413	P
CAL ICV CCV		X	1/26	0743	100	7/78	2408888			97	F
CAL ICV CCV		MM	1/27	1400	1413	10/4/17	KC1 K-8			1432	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL ICV CCV		MM	1/26	0730	7	4/18	2408804			70	P
CAL ICV CCV					4	8/18	2408803			40	F
CAL ICV CCV		X	1/26	0733	10	10/17	1408346			140	F
CAL ICV CCV		X	1/26	0735	4	6/18	2408803			407	F
CAL ICV CCV		MM	1/27	1402	7	4/18	2408804			716	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

Maintenance: Weekly pH Slope:

Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

XCB optical probe not functioning

no reading

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification