

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box if there is qualified data on this page.

Project/Site:

2017-03-06-47

Date:

3/17/17

Meter #

DEP 132237

Temperature (Quarterly)

For Date of Last Temperature Verification see

in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL ICV CCV	MM	3/17/17	0736				8.6	22.9	91.4	8.575	P
CAL ICV CCV	MM	3/8/17	1006				8.9	22.5	103	8.66	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

CAL ICV CCV	MM	3/17	0735		1413	10/4/17	KC16-8			1413	P
CAL ICV CCV	MM	3/17	0738		100	7/18	210888			169	P
CAL ICV CCV	MM	3/8	1005		58288	10/4/17	KC16-7			58288	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL ICV CCV	MM	3/17	0740		7	4/18	2108104			7.0	P
CAL ICV CCV	MM	3/17	0741		4	8/18	2108103			4.0	P
CAL ICV CCV	MM	3/17	0742		10	10/17	2108344			10.0	P
CAL ICV CCV	MM	3/17	0743		7	4/18	2108304			7.19	P
CAL ICV CCV	MM	3/8	1001		10	10/17	1608546			10.1	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

Maintenance: Weekly pH Slope:

Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: