


Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05



Project/Site:

Date: 3/15/17

Meter # DEP 132237

Temperature (Quarterly)

For Date of Last Temperature Verification see

In log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	NM	3/15	0715		9.3	20.1	103	9.074	(P) F
CAL	ICV	CCV									F
CAL	ICV	CCV	ES	3/15	1415		9.76	17.6	102.4	9.55	P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard umhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading umhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	NM	3/15	0720	M13	10/4/17		141.3	164	(P) F
CAL	ICV	CCV									F
CAL	ICV	CCV	ES	3/15	1412	58640	10/4/17		59755	96	(P) F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	NM	3/15	0730	7			-46.6	7.0	(P) F
CAL	ICV	CCV									F
CAL	ICV	CCV	ES	3/15	1413	4			4.0	4.0	(P) F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

Maintenance: Weekly pH Slope:

Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification