

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: DNV (KQ-2017-03-13-28) Date: 3/22/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see 2/16/17 in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Change	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL ICV CCV	PR	PR	3/22/17	080910	PR 3/22/17		8.8	22.4	102	8.7	P
CAL ICV CCV	PR	PR	3/23/17	0706			9.1	22.0	104	8.7	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μmhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL ICV CCV	PR	PR	3/22/17	0847	58600	10/4/17	KC1 16-7			58248 → 58600	P
CAL ICV CCV	PR	PR	3/22/17	0902	58640	10/4/17	KC1 16-7			58605	P
CAL ICV CCV	PR	PR	3/23/17	0648	12900	10/4/17	KC1 16-6			13127	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL ICV CCV	PR	PR	3/22/17	0851	7.0	Apr 2015	2605804			7.049 → 7.00	P
CAL ICV CCV	PR	PR	3/22/17		4.0	Aug 2015	2608503			3.973 → 4.00	P
CAL ICV CCV	PR	PR	3/22/17		10.0	Oct 2017	1605846			9.993 → 9.99	P
CAL ICV CCV	PR	PR	3/22/17	0900	10.0	Oct 2017	1605846			10.000 → 10.00	P
CAL ICV CCV	PR	PR	3/23/17	0637	4.0	Aug 2015	2608503			4.1	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No