

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project Site: 2017-04-10-30

Date: 4/10/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	SS	4/10/17	0813		8.68	22.6	100.6	8.644	P
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

CAL	ICV	CCV	SS	4/10/17	0818	58640	10/4/17	Kcl 16-7		58640	P
CAL	ICV	CCV	SS	4/10/17	0821	1413	10/4/17	Kcl 16-8		1337	P
CAL	ICV	CCV	SS	4/10/17	0824	58640	10/4/17	Kcl 16-7		58640	P
CAL	ICV	CCV	SS	4/10/17	0826	1413	10/4/17	Kcl 16-8		1421	P
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	SS	4/10/17	0830	7	4/10/17	2605604	-48.0	7	P
CAL	ICV	CCV	SS	4/10/17	0833	4	8/20/18	2608503	1365	4	P
CAL	ICV	CCV	SS	4/10/17	0838	10	10/20/17	1605B46	-211	10	P
CAL	ICV	CCV	SS	4/10/17	0837	4	8/20/18	2608503	#1202	4.1	P
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F
CAL	ICV	CCV									F

Maintenance: Weekly pH Slope: 174.5 Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: