

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: FD-2017-04-17-25

Date: 4/18/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Change	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	<u>W</u>	<u>4/18/17</u>	<u>0715</u>		<u>8.8</u>	<u>21.8</u>	<u>101</u>	<u>8.772</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/19/17</u>	<u>0730</u>		<u>8.7</u>	<u>22.1</u>	<u>101</u>	<u>8.595</u>	<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

CAL	ICV	CCV	<u>W</u>	<u>4/18</u>	<u>0720</u>	<u>12900</u>	<u>KCL 16-6</u>			<u>12900</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/18</u>	<u>0721</u>	<u>1413</u>	<u>KCL 16-8</u>			<u>1414</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/19</u>	<u>0735</u>	<u>100</u>	<u>260858</u>			<u>103</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/19</u>	<u>0737</u>	<u>58640</u>	<u>KCL 16-7</u>			<u>58622</u>	<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	<u>W</u>	<u>4/18</u>	<u>0725</u>	<u>7</u>	<u>2605864</u>		<u>-0.5</u>	<u>7.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/18</u>	<u>0726</u>	<u>4</u>	<u>2608503</u>		<u>-171.5</u>	<u>4.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/19</u>	<u>0728</u>	<u>10</u>	<u>2608544</u>		<u>-172.1</u>	<u>10.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/19</u>	<u>0729</u>	<u>4</u>	<u>2608503</u>		<u>-170.5</u>	<u>4.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>4/19</u>	<u>0740</u>	<u>10</u>	<u>1605846</u>		<u>-170.5</u>	<u>10.0</u>	<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>
CAL	ICV	CCV									<u>P</u>

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

changed pH probe