

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project Site: 267-66-05-63

Date: 6/5/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL ICV CCV		MM	6/6/5	0715			8.15	26.2	101	8.07	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV		FB	6/6/5	16:55			25.2	98	8.0		P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μ mhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ mhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------	-----------	-------	----------	---------------	-----------------------	--------------

CAL ICV CCV		MM	6/5	0717	1413	10/17	KC16-8			1413	P
CAL ICV CCV		Y	Y	0718	100	7/18	2108008			95	P
CAL ICV CCV											F
CAL ICV CCV		FB	6/5	16:59	100	7/18	2208888			102	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL ICV CCV		MM	6/5	0720	7	4/18	240804			7.0	P
CAL ICV CCV		Y	Y	0723	10	6/18	2108003			4.0	P
CAL ICV CCV		Y	Y	0725	4	6/18	2108003			3.9	P
CAL ICV CCV											F
CAL ICV CCV		FB	6/5	17:00	4	4/17	2008004			7.0	P
CAL ICV CCV											F
CAL ICV CCV											F
CAL ICV CCV											F

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification