

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: RD-2017-07-17-34

Date: 7/19/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	MM	7/19/17	0810		8.51	23.9	98	8.934	P
CAL	ICV	CCV	MM	7/20/17	0720		8.4	24.7	101	8.301	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard umhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading umhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	MM	7/19/17	0812	1413	10/4/17	KC17-1		1413	P
CAL	ICV	CCV	MM	7/20/17	0815	68490	10/4/17	KC17-1		10915	P
CAL	ICV	CCV	MM	7/20/17	0825	100	6/12/18	260888		104	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	MM	7/19/17	0820	7	4/18	20804		7.0	P
CAL	ICV	CCV	MM	7/19/17	0822	4	8/18	20807		4.0	P
CAL	ICV	CCV	MM	7/20/17	0824	10	4/17	108346		10.60	P
CAL	ICV	CCV	MM	7/20/17	0824	4	4/17	108346		10.04	P
CAL	ICV	CCV	MM	7/20/17	0730	4	8/18	20807		4.0	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No