

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box if there is qualified data on this page.

Project/Site: 20-2017-07-24-17

Date: 7/26/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	-----------------	--------------

Acceptance Criteria: $\pm 0.3 \text{ mg/L}$ (from chart)

CAL	ICV	CCV	<u>W</u>	<u>7/26</u>	<u>0730</u>		<u>8.62</u>	<u>22.8</u>	<u>100.0</u>	<u>8.611</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>7/27</u>	<u>0710</u>		<u>8.6</u>	<u>23.3</u>	<u>101</u>	<u>8.530</u>	<u>P</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

Acceptance Criteria: $\pm 5\%$

CAL	ICV	CCV	<u>W</u>	<u>7/26</u>	<u>0735</u>	<u>12900</u>	<u>6/6/18</u>	<u>KCE 172</u>		<u>12900</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>7/26</u>	<u>0736</u>	<u>1413</u>	<u>10/14/17</u>	<u>12C17-1</u>		<u>1410</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>7/27</u>	<u>0715</u>	<u>100</u>	<u>07/20/18</u>	<u>240888</u>		<u>103</u>	<u>P</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

Acceptance Criteria: $\pm 0.2 \text{ SU}$

CAL	ICV	CCV	<u>W</u>	<u>7/26</u>	<u>0740</u>	<u>7</u>	<u>04/20/18</u>	<u>2408804</u>	<u>-5.7</u>	<u>7.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>7/26</u>	<u>0741</u>	<u>4</u>	<u>08/20/18</u>	<u>2408803</u>	<u>167.0</u>	<u>4.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>7/26</u>	<u>0742</u>	<u>10</u>	<u>10/20/17</u>	<u>2408846</u>	<u>-179.1</u>	<u>10.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>7/27</u>	<u>0745</u>	<u>4</u>	<u>08/20/18</u>	<u>2408803</u>	<u>4.0</u>	<u>4.0</u>	<u>P</u>
CAL	ICV	CCV	<u>W</u>	<u>7/27</u>	<u>0720</u>	<u>10</u>	<u>10/20/17</u>	<u>2408846</u>	<u>10.0</u>	<u>10.0</u>	<u>P</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>
CAL	ICV	CCV									<u>F</u>

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: