

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: 202017-08-07-27

Date: 8/8/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	W	8/8	0715		8.27	25.6	101	8.173	P
CAL	ICV	CCV	W	8/9	0720		8.3	24.4	100	8.356	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

CAL	ICV	CCV	W	8/8	0720	12900	6/6/18	KCL 17-2		12900	P
CAL	ICV	CCV	W	8/9	0721	1473	10/4/17	KCL 17-1		1409	P
CAL	ICV	CCV	W	8/9	0725	100	02/20/19	2702495		104	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	W	8/8	0725	7	04/20/18	2608804	-11.4	7.0	P
CAL	ICV	CCV	W	8/9	0726	4	08/20/18	2608803	164.5	4.0	P
CAL	ICV	CCV	W	8/9	0727	10	10/20/17	1405346	-152.0	10.0	P
CAL	ICV	CCV	W	8/9	0730	4	08/20/18	2608803	4.0	4.0	P
CAL	ICV	CCV	W	8/9	0727	10	10/20/17	1405346	10.0	10.0	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: