

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: Reference Streams Turkey Cr Date: 8/20/17

Meter # DEP 132237

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

<u>ICV</u>	<u>CCV</u>	<u>SS</u>	<u>8/20/17</u>	<u>0841</u>			<u>8.66</u>	<u>23.2</u>	<u>101</u>	<u>8.514</u>	<u>P</u>
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

<u>ICV</u>	<u>CCV</u>	<u>SS</u>	<u>8/20/17</u>	<u>0844</u>	<u>12900</u>	<u>6/6/18</u>	<u>KCL 17-2</u>			<u>12900</u>	<u>P</u>
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

<u>ICV</u>	<u>CCV</u>	<u>SS</u>	<u>8/20/17</u>	<u>0849</u>	<u>7</u>	<u>4/20/18</u>	<u>2605804</u>		<u>-11.2</u>	<u>7.0</u>	<u>P</u>
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

did not need the test this day

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification