

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: 2Q-2017-01-09-47 Date: 01/17/17 Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
Dissolved Oxygen										
CAL ICV CCV	<u>W</u>	<u>1/17/17</u>	<u>0715</u>			<u>8.40</u>	<u>23.8</u>	<u>99.4</u>	<u>8.40</u>	<u>P</u>
CAL ICV CCV	<u>W</u>	<u>1/18/17</u>	<u>1350</u>			<u>8.17</u>	<u>22.4</u>	<u>101</u>	<u>8.677</u>	<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>

DEP SOP FT 1200	Initials	Date	Time	Standard µmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading µmhos/cm	Pass or Fail
Specific Conductance										
CAL ICV CCV	<u>W</u>	<u>1/17/17</u>	<u>0720</u>	<u>15900</u>	<u>10/14/17</u>	<u>KCL 16-4</u>			<u>12900</u>	<u>P</u>
CAL ICV CCV	<u>W</u>	<u>1/17/17</u>	<u>0721</u>	<u>1417</u>	<u>10/14/17</u>	<u>KCL 16-8</u>			<u>1417</u>	<u>P</u>
CAL ICV CCV	<u>W</u>	<u>1/18/17</u>	<u>1355</u>	<u>100</u>	<u>07/20/18</u>	<u>2608888</u>			<u>105</u>	<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>

DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
pH										
CAL ICV CCV	<u>W</u>	<u>1/17/17</u>	<u>0725</u>	<u>7</u>	<u>04/18</u>	<u>2605804</u>		<u>-29.5</u>	<u>7.0</u>	<u>P</u>
CAL ICV CCV	<u>W</u>	<u>1/18/17</u>	<u>0726</u>	<u>4</u>	<u>08/20/17</u>	<u>2608F03</u>		<u>146.8</u>	<u>4.0</u>	<u>P</u>
CAL ICV CCV	<u>W</u>	<u>1/18/17</u>	<u>0727</u>	<u>10</u>	<u>10/20/17</u>	<u>1605B46</u>		<u>-19.1</u>	<u>10.0</u>	<u>P</u>
CAL ICV CCV	<u>W</u>	<u>1/18/17</u>	<u>0728</u>	<u>4</u>	<u>08/20/17</u>	<u>2608F03</u>			<u>4.0</u>	<u>P</u>
CAL ICV CCV	<u>W</u>	<u>1/18/17</u>	<u>1400</u>	<u>10</u>	<u>10/20/17</u>	<u>1605B46</u>			<u>10.0</u>	<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>
CAL ICV CCV										<u>P</u>

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No