

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project Site: SMP/120-2017-01-23-35 Est Date: 02/01/2017 Eglia West Est/NA Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see In log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	-----------------	--------------

CAL	ICV	CCV	FB	2/01	8:44		8.94	22.0	100.1	8.98	P
CAL	ICV	CCV	W	2/1	0:25		8.75	22.0	100.1	8.94	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μmhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	FB	2/01	8:45	1413	10/4/17	KCL16-8	1	1413	P
CAL	ICV	CCV	W	2/1	8:50	100	1/7/18	2608088	1	98	P
CAL	ICV	CCV			0:20	100	7/12	2608088	1	101	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	FB	2/01	8:55	7.4	4/20/17	2605804	-33.2	7.4	P
CAL	ICV	CCV			9:00	10.0	10/20/17	1605346	-10.3	9.99	P
CAL	ICV	CCV			9:05	4.0	2/20/18	2608103	144.3	4.00	P
CAL	ICV	CCV	W	2/1	9:10	7.0	4/20/18	2608103	7	7.05	P
CAL	ICV	CCV			0:25	4.0				4.0	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed? Yes No