

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site:

PR 2017-02-13-26

Date:

2/13/17

Meter # **DEP 113529**

Temperature (Quarterly) For Date of Last Temperature Verification see _____ in log book

Dissolved Oxygen

DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV CCV	MM	2/13	0715			8.62	22.9	100	8.515	P
CAL ICV CCV	MM	2/14	1530			8.7	2.8	99	8.777	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Specific Conductance

DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL ICV CCV	MM	2/13	0719	1413	10/4/17	16116-8			1413	P
CAL ICV CCV	J	2/14	0721	100	7/18	2608888			100	P
CAL ICV CCV	MM	2/14	1535	58640	10/4/17	16116-7			58912	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

pH

DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL ICV CCV	MM	2/13	0720	7	4/18	260804			7.0	P
CAL ICV CCV	J	2/14	0721	9	8/18	2608703			4.0	P
CAL ICV CCV	J	2/14	0723	10	8/18	2608396			10.0	P
CAL ICV CCV	MM	2/14	1540	10	8/18	2608703			10.03	P
CAL ICV CCV					10/17	1605746			10.01	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Maintenance: Weekly pH Slope: _____

Notes:

Specific Conductance Probe Cleaned? Yes No

Dissolved Oxygen Membrane Changed: Yes No