

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: RQ-2017-03-13-27

Date: 3/15/17

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	SS	3/15/17	0810		9.2	20.1	101	9.074	P
CAL	ICV	CCV	SS	3/16/17	0918		9.2	19.6	101	9.165	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard umhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading umhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	SS	3/15/17	0813	12,900	10/4/17	Kel 16-6		12,900	P
CAL	ICV	CCV	SS	3/15/17	0814	1413	10/29/17	Kel 16-8		1437	P
CAL	ICV	CCV	SS	3/16/17	0921	100	07/18	260588		98	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	SS	3/15/17	0817	7	4/18	2605804	-33.6	7	P
CAL	ICV	CCV	SS	3/16/17	0819	4	8/18	2608503	142.0	4	P
CAL	ICV	CCV	SS	3/16/17	0821	10	10/17	1605846	-205.6	10.0	P
CAL	ICV	CCV	SS	3/16/17	0825	24	8/18	2608503	139.7	4.0	P
CAL	ICV	CCV	SS	3/16/17	0925	7	4/18	2605804	-35.6	7.1	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes (No) Dissolved Oxygen Membrane Changed: Yes (No)