

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: 20-2017-04-03-72 Date: 4/13/17 Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0800</u>			<u>8.9</u>	<u>21.6</u>		<u>8.812</u>	<u>P</u>
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0400</u>			<u>8.5</u>	<u>23.0</u>		<u>8.578</u>	<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μ mhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ mhos/cm	Pass or Fail
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0805</u>	<u>12900</u>	<u>10/4/17</u>	<u>KCL 16-0</u>			<u>12900</u>	<u>P</u>
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0807</u>	<u>1413</u>	<u>10/4/17</u>	<u>KCL 16-0</u>			<u>1414</u>	<u>P</u>
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0410</u>	<u>100</u>	<u>07/20/18</u>	<u>260888</u>			<u>101</u>	<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0810</u>	<u>7</u>	<u>05/20/18</u>	<u>2605804</u>			<u>7.0</u>	<u>P</u>
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0811</u>	<u>4</u>	<u>08/20/18</u>	<u>2608503</u>			<u>4.0</u>	<u>P</u>
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0812</u>	<u>10</u>	<u>10/20/17</u>	<u>1605346</u>			<u>10.0</u>	<u>P</u>
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0813</u>	<u>4</u>	<u>08/20/18</u>	<u>2608503</u>			<u>4.0</u>	<u>P</u>
CAL ICV CCV		<u>N</u>	<u>4/13/17</u>	<u>0415</u>	<u>10</u>	<u>10/20/17</u>	<u>1605346</u>			<u>10.0</u>	<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>
CAL ICV CCV											<u>P</u>

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No