

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: RA-201704-10-71

Date: 4/17/17

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL ICV <u>CCV</u>		<u>MM</u>	<u>4/17/17</u>	<u>0715</u>			<u>8.408</u>	<u>24.0</u>	<u>106</u>	<u>8.418</u>	<u>(P)</u>
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>		<u>MM</u>	<u>4/17/17</u>	<u>1600</u>			<u>8.64</u>	<u>23.0</u>	<u>161</u>	<u>8.578</u>	<u>(P)</u>
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

CAL ICV <u>CCV</u>		<u>MM</u>	<u>4/17/17</u>	<u>0720</u>	<u>1413</u>	<u>10/4/17</u>	<u>KC14-8</u>			<u>1413</u>	<u>(P)</u>
CAL ICV <u>CCV</u>		<u>Y</u>	<u>4/17/17</u>	<u>0722</u>	<u>100</u>	<u>7/18</u>	<u>2108888</u>			<u>95</u>	<u>(P)</u>
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>		<u>MM</u>	<u>4/17/17</u>	<u>1602</u>	<u>100</u>	<u>7/18</u>	<u>2108888</u>			<u>105</u>	<u>(P)</u>
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL ICV <u>CCV</u>		<u>MM</u>	<u>4/17/17</u>	<u>0727</u>	<u>7</u>	<u>4/18</u>	<u>2108804</u>			<u>7.0</u>	<u>(P)</u>
CAL ICV <u>CCV</u>		<u>Y</u>	<u>4/17/17</u>	<u>0728</u>	<u>4</u>	<u>8/18</u>	<u>2108503</u>			<u>4.0</u>	<u>(P)</u>
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>		<u>Y</u>	<u>4/17/17</u>	<u>0735</u>	<u>4</u>	<u>8/18</u>	<u>2108503</u>			<u>1357</u>	<u>(P)</u>
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>		<u>MM</u>	<u>4/17/17</u>	<u>1600</u>	<u>7</u>	<u>4/18</u>	<u>2108804</u>				<u>(P)</u>
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F
CAL ICV <u>CCV</u>											F

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: