

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: 2017-04-10-32

Date: 4/18/17

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	NM	4/18/17	0800			8.64	23.0	101	8.58	P
CAL	ICV	CCV	NM	4/19/17	0730			8.18	22.9	101	8.55	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard μ mhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ mhos/cm	Pass or Fail
CAL	ICV	CCV	NM	4/18/17	0805	1413	6/4/17	KC16-8			1413	P
CAL	ICV	CCV	NM	4/19/17	0807	100	7/18	2408888			100	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	NM	4/18/17	0810	7	2408888	4/18			7.0	P
CAL	ICV	CCV	NM	4/19/17	0811	4	2408888	8/18			4.0	P
CAL	ICV	CCV	NM	4/19/17	0812	20	2408888	8/17			4.0	P
CAL	ICV	CCV	NM	4/19/17	0815	4	2408888	8/18			4.5	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No