

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Meter # **DEP 113529**

Date: **6/5/17**

Project/Site: **2017-05-08-42**

Temperature (Quarterly) For Date of Last Temperature Verification see _____ in log book									
DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)
Dissolved Oxygen									

CAL ICV CCV	JS	6/5/17	0750			8.05	26.1	99	8.099	P
CAL ICV CCV	JS	6/6/17	0925			8.52	23.1	100	8.562	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

DEP SOP FT 1200	Initials	Date	Time	Standard µmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading µmhos/cm	Pass or Fail
Specific Conductance										

CAL ICV CCV	JS	6/5/17	0754	1413	10/4/17	KCL16-8			1413	P
CAL ICV CCV	JS	6/5/17	0756	100	7/14	2608888			99	P
CAL ICV CCV	JS	6/6/17	0927	100	7/18	2608888			102	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
pH										

CAL ICV CCV	JS	6/5/17	0757	7	4/18	2605804		-41.5	7.00	P
CAL ICV CCV	JS	6/5/17	0759	4	8/18	2608503		133.6	4.00	P
CAL ICV CCV	JS	6/5/17	0801	10	10/17	1605846		-212.1	9.99	P
CAL ICV CCV	JS	6/5/17	0803	4	8/18	2608503		131.2	4.02	P
CAL ICV CCV	JS	6/6/17	0929	2	4/18	2608504		-43.1	7.05	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Maintenance: Weekly pH Slope: **175.1** Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: **71.6**