

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: 2017-06-19-67

Date: 6/26/17

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Change	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	-----------------	--------------

CAL	ICV	CCV	SS	6/26/17	0804		8.13	26.0	106	8.114	P
CAL	ICV	CCV	SS	6/27/17	0842		8.57	23.5	101	8.418	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μmhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	SS	6/26/17	0807	12900	6/6/18	Kel 17-2		12900	P
CAL	ICV	CCV	SS	6/26/17	0810	100	10/10/18	200 8889		105	P
CAL	ICV	CCV	SS	6/27/17	0850	1413	10/4/17	Kel 17-1		1448	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	SS	6/26/17	0812	7	4/18	2605804	-49.9	7.0	P
CAL	ICV	CCV	SS	6/26/17	0814	4	8/18	2606503	127.2	4.0	P
CAL	ICV	CCV	SS	6/26/17	0817	10	10/17	1655346	-217.4	40.0	P
CAL	ICV	CCV	SS	6/26/17	0819	4	8/18	2605503	127	4.0	P
CAL	ICV	CCV	SS	6/27/17	0853	7	4/18	2605804	-472.0	2.01	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed? Yes No