

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: 2017-07-10-34

Date: 7/11/17

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	MM	7/11	0730		8.28	25.3	101	8.38	P
CAL	ICV	CCV	SS	7/12	1320		8.73	22.4	101	8.07	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μmhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	MM	7/11	0732	1413	10447	KC117-1		1413	P
CAL	ICV	CCV	SS	7/12	0734	100	718	20088		107	P
CAL	ICV	CCV			1330	100	718	20088		96	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL	ICV	CCV	MM	7/11	0740	7	20088	4116		7.0	P
CAL	ICV	CCV			0742	4	818	20088		4.0	P
CAL	ICV	CCV			0744	10	10084			10.0	P
CAL	ICV	CCV			0745	10	10084			10.33	P
CAL	ICV	CCV	SS	7/12	1335	4	911	20088		4.10	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No