

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site: LL11015 Deer Pt Date: 7/14/17 Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen

DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV CCV	SS	7/14/17	0805				8.51	9.434	P
CAL ICV CCV									P
CAL ICV CCV									P
CAL ICV CCV									P
CAL ICV CCV									P
CAL ICV CCV									P
CAL ICV CCV									P
CAL ICV CCV									P

Specific Conductance

DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL ICV CCV	SS	7/14/17	0807	1413	10/4/17	Kel 17-1			1413	P
CAL ICV CCV	SS	7/14/17	0811	100	7/18	2608884			101	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

pH

DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
CAL ICV CCV	SS	7/14/17	0814	7	4/8/2018	2605804	10	-46.2	7.00	P
CAL ICV CCV	SS	7/14/17	0815	4	8/18	2608503		131.5	4.00	P
CAL ICV CCV	SS	7/14/17	0817	10	10/17	1605846		-205.9	9.95	P
CAL ICV CCV	SS	7/14/17	0819	4	8/18	2608503		133.1	3.9	P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P
CAL ICV CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:

Pid not sample
due to rain

Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification