

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box if there is qualified data on this page.

Project/Site: 207-08-21-17

Date: 8/2/17

Meter # DEP 113529

For Date of Last Temperature Verification see _____ in log book									
Temperature (Quarterly)	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO
Dissolved Oxygen									

CAL ICV CCV	8/2/17	0805					8.36	24.6	100	Pass or Fail
CAL ICV CCV	8/3/17	1343					8.58	23.0	100	Pass or Fail
CAL ICV CCV										Pass or Fail
CAL ICV CCV										Pass or Fail
CAL ICV CCV										Pass or Fail
CAL ICV CCV										Pass or Fail
CAL ICV CCV										Pass or Fail
CAL ICV CCV										Pass or Fail

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μ mhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μ mhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------	-----------	-------	----------	---------------	-----------------------	--------------

CAL ICV CCV	8/2/17	0808			12,900	6/6/18	Kel 12-2			12,900	Pass or Fail
CAL ICV CCV	8/2/17	0812			100	6/7/18	2608888			102	Pass or Fail
CAL ICV CCV	8/3/17	1348			100	6/7/18	2608888			102	Pass or Fail
CAL ICV CCV											Pass or Fail
CAL ICV CCV											Pass or Fail
CAL ICV CCV											Pass or Fail
CAL ICV CCV											Pass or Fail
CAL ICV CCV											Pass or Fail

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	mV / Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	------------	------------	--------------

CAL ICV CCV	8/2/17	0814			7	4/20/18	2605804			7.0	Pass or Fail
CAL ICV CCV		0816			4	8/20/18	2608803			4.0	Pass or Fail
CAL ICV CCV		0817			10	10/17	1605846			9.97	Pass or Fail
CAL ICV CCV		0819			4	8/20/18	2608803			3.97	Pass or Fail
CAL ICV CCV		1350			7	4/20/18	2605800			7.03	Pass or Fail
CAL ICV CCV											Pass or Fail
CAL ICV CCV											Pass or Fail
CAL ICV CCV											Pass or Fail
CAL ICV CCV											Pass or Fail

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No