

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project Site: 9/6/17 2017-09-04-46 Date: 9/6/17 Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	-----------------	----------	------	------	--------------	------------	------	---------	------	------------------------------	--------------

CAL	ICV	CCV	88	9/6/17	0925		8.67	22.5	100	8.660	P
CAL	ICV	CCV	88	9/31/17	0904		8.64	22.6	100	8.644	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard μmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading μmhos/cm	Pass or Fail
----------------------	-----------------	----------	------	------	-------------------	-----------	-------	----------	---------------	------------------	--------------

CAL	ICV	CCV	88	9/6/17	0930	12900	6/6/18			12900	P
CAL	ICV	CCV	88	9/6/17	0932	1413	10/4/17			1424	P
CAL	ICV	CCV	88	9/7/17	0906	106	02/20/19			103	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	-------	------------	--------------

CAL	ICV	CCV	88	9/6/17	0934	7.0	4/20/18			-50.7	7.0	P
CAL	ICV	CCV			0936	4.0	10/20/18			126.1	4.0	P
CAL	ICV	CCV			0938	10.0	10/20/17			-216.5	9.97	P
CAL	ICV	CCV			0940	4.0	10/20/18			124.1	4.0	P
CAL	ICV	CCV			0908	2.0	4/20/18			-51.5	7.08	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No