

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Meter # **DEP 113529**

Project/Site: RD-2017-09-25-19

Date: 9/25/17

In log book

For Date of Last Temperature Verification see

Dissolved Oxygen		DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL	ICV	CCV	MM	9/25/17	08:55			23.9	100	8.934	P
CAL	ICV	CCV									P
CAL	ICV	CCV	FB	9/25/17	16:00			22.8	100	8.61	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance		DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL	ICV	CCV	MM	9/25/17	08:20	1413	9/20/18	141174			1413	P
CAL	ICV	CCV	J	9/25/17	08:23	100	2/19	200006			102	P
CAL	ICV	CCV	FS	9/25/17	10:05	1413	9/20/18	141174			1420	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

pH		DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail
CAL	ICV	CCV	MM	9/25/17	08:25	7	4/18	260504			7.0	P
CAL	ICV	CCV	J	9/25/17	08:27	4	10/18	260504			4.0	P
CAL	ICV	CCV	J	9/25/17	08:31	10	10/17	141584			14.0	P
CAL	ICV	CCV	FB	9/25/17	16:00	4	10/18	260504		125.3	3.83	P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P
CAL	ICV	CCV										P

Maintenance: Weekly pH Slope: _____

Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes: