

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Boldly "X" this box
if there is qualified
data on this page.

Project/Site: R2-2017-09-28-23

Date: 9/20/2017

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
------------------	--------------------	----------	------	------	-----------------	---------------	------	---------	------	------------------------------------	-----------------

CAL	ICV	CCV	FB	9/20/17	9:15		8.5	23.6	100	9.43	P
CAL	ICV	CCV	FB	9/20/17	9:00		8.8	21.7	100	9.83	P
CAL	ICV	CCV	FB	9/20/17	9:00		8.8	21.7	100	9.83	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard µmhos/cm	Exp. Date	Lot #	Bottle #	Cell Constant	Reading µmhos/cm	Pass or Fail
----------------------	--------------------	----------	------	------	----------------------	-----------	-------	----------	------------------	---------------------	-----------------

CAL	ICV	CCV	FB	9/20/17	9:33		1413	9/20/17	KEC17-Y		P
CAL	ICV	CCV	FB	9/20/17	9:40		100	02/19	2703996		P
CAL	ICV	CCV	FB	9/20/17	9:05		1413	9/20/17	KEC17-Y		P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail
----	--------------------	----------	------	------	----------------	-----------	-------	----------	-------	---------------	-----------------

CAL	ICV	CCV	FB	9/20/17	9:20		7	4/18	2605804		P
CAL	ICV	CCV	FB		9:22		4	10/18	2611883		P
CAL	ICV	CCV	FB		9:25		10	10/17	1605846		P
CAL	ICV	CCV	FB				4	10/18	2611883		P
CAL	ICV	CCV	FB		9:30		7	4/18	260		P
CAL	ICV	CCV	FB		9:10		7	4/18	260		P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: _____ Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No