

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Project/Site:

2017-09-11-31

Date:

10/2/17

Meter # DEP 113529

Temperature (Quarterly) For Date of Last Temperature Verification see in log book

Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge	Probe Gain	mg/L	Temp °C	% DO	Saturation mg/L (from chart)	Pass or Fail
CAL ICV CCV		NM	10/2	0731			8.31	25.2	101	8.233	P
CAL ICV CCV		SS	10/2	1005			8.75	23.06	100.3	8.727	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
CAL ICV CCV		NM	10/2	0735	1413	6/6/18	NCL 17-2			1413	P
CAL ICV CCV		SS	10/2	0737	100	2/19	2000000			104	P
CAL ICV CCV			10/2	1012	1413	6/6/18	NCL 17-2			1412	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail
CAL ICV CCV		NM	10/2	0740	7	4/18	2605804			7.0	P
CAL ICV CCV		SS	10/2	0742	4	2/18	2411883			4.0	P
CAL ICV CCV			10/2	0744	10	10/17	1608346			10.0	P
CAL ICV CCV		SS	10/2	0745	4	10/18	2601883			3.87	P
CAL ICV CCV			10/2	1007	7	4/18	2605804			7.09	P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P
CAL ICV CCV											P

Maintenance: Weekly pH Slope: 18.2 Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No