

Boldly "X" this box
if there is qualified
data on this page.

Form FD9000-8 CALIBRATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000) 11-10-05

Meter # **DEP 113529**

Date: **10/5/17**

Project/Site: **2017-10-02-40**

Temperature (Quarterly)		For Date of Last Temperature Verification see		in log book	
Dissolved Oxygen	DEP SOP FT 1500	Initials	Date	Time	Probe Charge

Pass or Fail	Saturation mg/L (from chart)	% DO	Temp °C	mg/L	Probe Gain
F	101	100.0	23.3	8.65	
F			23.4	8.55	
F					
F					
F					
F					
F					
F					
F					
F					

Specific Conductance	DEP SOP FT 1200	Initials	Date	Time	Standard $\mu\text{mhos/cm}$	Exp. Date	Lot #	Bottle #	Cell Constant	Reading $\mu\text{mhos/cm}$	Pass or Fail
----------------------	-----------------	----------	------	------	------------------------------	-----------	-------	----------	---------------	-----------------------------	--------------

CAL	ICV	CCV	10/5	0805	1413	10/18	KC1172			1413	P
CAL	ICV	CCV	10/6	0808	100	2/19	2703996			103	P
CAL	ICV	CCV	10/6	1405	100	2/12	2703996			98	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

pH	DEP SOP FT 1100	Initials	Date	Time	Standard SU	Exp. Date	Lot #	Bottle #	Slope	Reading SU	Pass or Fail
----	-----------------	----------	------	------	-------------	-----------	-------	----------	-------	------------	--------------

CAL	ICV	CCV	10/5	0810	7	4/18	2608804			7.0	P
CAL	ICV	CCV	10/6	0812	4	10/18	2611883			4.0	P
CAL	ICV	CCV	10/6	0814	10	10/17	11653246			10.0	P
CAL	ICV	CCV	10/6	0817	4	10/18	2611883			122.7	P
CAL	ICV	CCV	10/6	1403	4	4/18	2608804			3.2	P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P
CAL	ICV	CCV									P

Maintenance: Weekly pH Slope: **1.22** Specific Conductance Probe Cleaned? Yes No Dissolved Oxygen Membrane Changed: Yes No

Notes:
Perform only in Calibrate Mode: CAL - Calibrate
Perform only in Run Mode: ICV - Initial Calibration Verification
Perform only in Run Mode: CCV - Continuing Calibration Verification