

Biological Analysis Report

NW-DIST-2017-10-20-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Overflow Analyses Performed By:
University of West Florida
11000 University Parkway
Pensacola, FL 32514
DOH Accreditation E71969

Event Description: **ENR D WBID 548GB 738 429**
Request ID: **RQ-2017-10-16-23**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 Government Center
Suite 208A
Pensacola, FL 32502-5794
Attn: Michael King

For additional information please contact
Cheryl Swanson - Administrator
J. Marshall Faircloth - Aquatic Toxicology
Jacqueline Savage - Bench Biology
Jacqueline Savage - Microbiology
Puja Jasrotia, Ph.D. - Molecular Biology
Jennifer Walters - Taxonomy (Invertebrate & Algal)
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8177

Certified by: Cheryl A. Swanson, Program Administrator

Date Certified: 22-NOV-2017 13:14



NON-CONFORMANCE REPORT INCLUDED

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Chlorophyll/Grain Size/BOD, Molecular and Overflow.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Dead River above Yellow River

Collection Date/Time: 10/19/2017 11:10

Field ID: G4NW0408

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939047	SM 10200 H (mod.)	Chlorophyll-a, Corrected	1.4	U	ug/L	P335013	
		Phaeophytin-a	1.0	U	ug/L	P335013	
		Chlorophyll-a, Uncorrected	1.2	IV	ug/L	P335013	

Ref. Method and Comment:

SM 10200 H (mod.): Precision data not available for at least 1 component due to sm amt analyte in QC sample. Refer to QA report for available precision data. Result V-qualified due to laboratory blank failure for all components. See NCR #6745.

Sample Location: Blackwater Bay west of Escribano Point

Collection Date/Time: 10/19/2017 12:35

Field ID: G4NW0401

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939043	SM 10200 H (mod.)	Chlorophyll-a, Corrected	1.5	IV	ug/L	P335013	
		Phaeophytin-a	0.40	U	ug/L	P335013	
		Chlorophyll-a, Uncorrected	1.6	V	ug/L	P335013	

Ref. Method and Comment:

SM 10200 H (mod.): Precision data not available for at least 1 component due to sm amt analyte in QC sample. Refer to QA report for available precision data. Result V-qualified due to laboratory blank failure for all components. See NCR #6745.

Sample Location: Blackwater Bay west of Escribano Point

Collection Date/Time: 10/19/2017 12:35

Field ID: G4NW0401

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939048	Enterolert/QT	Enterococci-Quanti-Tray	10		MPN/100 mL		
	SM 9222 D	Fecal Coliforms-Membrane Filter	15	B	cfu/100 mL		

Sample Location: Blackwater Bay west of Escribano Point

Collection Date/Time: 10/19/2017 12:35

Field ID: G4NW0401

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939050	SOP-PCR-4.0	HF183 - Human Source Marker**	Below Threshold				

Sample Location: Blackwater Bay west of Escribano Point

Collection Date/Time: 10/19/2017 12:40

Field ID: G4NW0401DUPLICATE

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939044	SM 10200 H (mod.)	Chlorophyll-a, Corrected	1.4	IV	ug/L	P335013	
		Phaeophytin-a	0.40	U	ug/L	P335013	
		Chlorophyll-a, Uncorrected	1.5	V	ug/L	P335013	

Ref. Method and Comment:

SM 10200 H (mod.): Precision data not available for at least 1 component due to sm amt analyte in QC sample. Refer to QA report for available precision data. Result V-qualified due to laboratory blank failure for all components. See NCR #6745.

Sample Location: Field Blank

Collection Date/Time: 10/19/2017 12:45

Field ID: Field Blank

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939045	SM 10200 H (mod.)	Chlorophyll-a, Corrected	0.55	U	ug/L	P335013	
		Phaeophytin-a	0.40	U	ug/L	P335013	

Field ID: Field Blank

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939045	SM 10200 H (mod.)	Chlorophyll-a, Uncorrected	0.40	U	ug/L	P335013	

Ref. Method and Comment:

SM 10200 H (mod.): Precision data not available for at least 1 component due to sm amt analyte in QC sample. Refer to QA report for available precision data. Result V-qualified due to laboratory blank failure for all components. See NCR #6745.

Sample Location: Texar Bayou

Collection Date/Time: 10/19/2017 14:05

Field ID: G4NW0402

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939046	SM 10200 H (mod.)	Chlorophyll-a, Corrected	7.0	AV	ug/L	P335013	
		Phaeophytin-a	0.44	U	ug/L	P335013	
		Chlorophyll-a, Uncorrected	7.3	AV	ug/L	P335013	

Ref. Method and Comment:

SM 10200 H (mod.): Laboratory duplicate result is 0.44 U for Phaeophytin-a. Precision data is unavailable due to the small amount of analyte in the QC sample. Result is V-qualified due to laboratory blank failure for all components. See NCR #6745.

Sample Location: Texar Bayou

Collection Date/Time: 10/19/2017 14:05

Field ID: G4NW0402

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939049	Enterolert/QT	Enterococci-Quanti-Tray	10		MPN/100 mL		

Sample Location: Texar Bayou

Collection Date/Time: 10/19/2017 14:05

Field ID: G4NW0402

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1939051	SOP-PCR-4.0	HF183 - Human Source Marker**	Below Threshold				

Non-Conformance Report

NCR ID: 6745

Event(s)

NW-DIST-2017-10-20-01
 NW-DIST-2017-10-20-01
 NW-DIST-2017-10-20-01
 NW-DIST-2017-10-20-01
 NW-DIST-2017-10-20-01

Job(s)

TLH-2017-10-20-50
 TLH-2017-10-20-50
 TLH-2017-10-20-50
 TLH-2017-10-20-50
 TLH-2017-10-20-50

Sample(s)

1939043
 1939044
 1939045
 1939046
 1939047

Test(s)

CHLSUITE-W
 CHLSUITE-W
 CHLSUITE-W
 CHLSUITE-W
 CHLSUITE-W

NCR Type: ANALYSIS

NCR Category: Analytical Measurement

Observation: Laboratory blank failure was observed in the analysis. Data values suggest the blank and another sample(s) were mislabeled, but this can't be verified.

Resolution: Affected batch uploaded per SOP BB-29 regarding blank failures. All affected samples were V qualified.

Authorized by/Date: Cheryl A. Swanson 11/11/2017

The Non-Conformance Report details exceptions or problems encountered with the events/jobs/samples/test.
 Please address questions to:

Chemistry Colin Wright (850) 245-8085
 Biology Cheryl Swanson (850) 245-8177

Quality Assurance Report Method Blank Results

Reference Method: SM 10200 H (mod.)
Batch ID: P335013

Component	Result	Code	Units
Chlorophyll-a, Corrected	5.7		ug/L
Chlorophyll-a, Uncorrected	6.3		ug/L
Phaeophytin-a	0.83	I	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 10200 H (mod.)
Batch ID: P335013

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlorophyll-a, Corrected	104		P	84 - 116

Quality Assurance Report Precision

Reference Method: SM 10200 H (mod.)
Batch ID: P335013

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1939046	Chlorophyll-a, Corrected	4.10	Sample	P	0 - 25
1939046	Chlorophyll-a, Uncorrected	6.58	Sample	P	0 - 25

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
SM 10200 H (mod.)	Chlorophyll-a, Corrected	104			4.10	
	Chlorophyll-a, Uncorrected				6.58	

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
Enterolert/QT / E71969	Enterococci by Enterolert Quanti-Tray method by UWF Overflow Laboratory for NWD	1939048, 1939049
SM 10200 H (mod.) / E31780	Phytoplankton chlorophyll-a corrected, uncorrected, and phaeophytin by spectrophotometry	1939043, 1939044, 1939045, 1939046, 1939047
SM 9222 D / E71969	Fecal coliforms by membrane filter method by UWF Overflow Laboratory for NWD	1939048
SOP-PCR-4.0 / E31780	Preparation of Samples by Filtration and held for qPCR Analysis	1939050, 1939051

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
Enterolert/QT	10/19/2017	10/19/2017 15:55		10/20/2017 16:25	UWF Lab	1939048, 1939049
SM 10200 H (mod.)	10/20/2017	10/20/2017 14:45	Carina A. Tull	10/31/2017	Kim M. Spencer	1939043, 1939044, 1939045, 1939046, 1939047
SM 9222 D	10/19/2017	10/19/2017 16:20		10/20/2017 14:42	UWF Lab	1939048
SOP-PCR-4.0	10/20/2017	10/20/2017 12:30	Sarah Menz	11/14/2017	Puja Jasrotia	1939050, 1939051