

Chemical Analysis Report

NW-DIST-2017-12-05-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Shoal River/Yellow WBID 160A 160B 30 BLANK DUP**
Request ID: **RQ-2017-11-27-13**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 W Government St
Suite 208A
Pensacola, FL 32502-5794
Attn: Michael King

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 22-DEC-2017 10:51

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: YELLOW RIVER

Collection Date/Time: 12/04/2017 10:45

Field ID: G4NW0370

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1950835	EPA 180.1 Rev. 2.0	Turbidity	5.1		NTU	P336430	
	EPA 300.0 Rev. 2.1	Chloride	3.3		mg Cl/L	P336534	
		Sulfate	0.99		mg SO4/L	P336536	
	SM 2120 B	Color (true)	34		PCU	P336417	
	SM 2320 B-97	Total Alkalinity	13		mg CaCO3/L	P336693	
	SM 2540 C-97	TDS	32		mg/L	P336955	
	SM 2540 D-97	TSS	6	I	mg/L	P337094	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P337104	
1950838	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P336935	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.21		mg N/L	P336754	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.21		mg N/L	P336498	
	EPA 365.1 Rev. 2.0	Total-P	0.011		mg P/L	P336732	
	SM 5310 B-00	Organic Carbon	2.3		mg C/L	P336636	
1950841	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P336405	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: DEAD RIVER ABOVE YELLOW RIVER

Collection Date/Time: 12/04/2017 12:15

Field ID: G4NW0408

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1950836	EPA 180.1 Rev. 2.0	Turbidity	6.0		NTU	P336430	
	EPA 300.0 Rev. 2.1	Chloride	4.6		mg Cl/L	P336534	
		Sulfate	0.81		mg SO4/L	P336536	
	SM 2120 B	Color (true)	38	A	PCU	P336417	
	SM 2320 B-97	Total Alkalinity	5.4		mg CaCO3/L	P336693	
	SM 2540 C-97	TDS	31		mg/L	P336955	
	SM 2540 D-97	TSS	3	I	mg/L	P337094	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P337104	
1950839	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P336935	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.27		mg N/L	P336754	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.28		mg N/L	P336498	
	EPA 365.1 Rev. 2.0	Total-P	0.014		mg P/L	P336732	
	SM 5310 B-00	Organic Carbon	2.6		mg C/L	P336636	
1950842	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P336405	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: ESCAMBIA RIVER UPSTREAM OF GULF POWER DISCHARGE

Collection Date/Time: 12/04/2017 13:50

Field ID: G4NW0264

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1950837	EPA 180.1 Rev. 2.0	Turbidity	6.2		NTU	P336430	
	EPA 300.0 Rev. 2.1	Chloride	1.2E+03		mg Cl/L	P336534	

Field ID: G4NW0264

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1950837	EPA 300.0 Rev. 2.1	Sulfate	170		mg SO4/L	P336536	
	SM 2120 B	Color (true)	27		PCU	P336417	
	SM 2320 B-97	Total Alkalinity	35		mg CaCO3/L	P336693	
	SM 2540 C-97	TDS	2.37E+03		mg/L	P336961	
	SM 2540 D-97	TSS	4	I	mg/L	P337097	
	SM 4500 F-C-97	Fluoride	0.12		mg F/L	P336697	
1950840	EPA 350.1 Rev. 2.0	Ammonia-N	0.058		mg N/L	P336935	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.28		mg N/L	P336754	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.35		mg N/L	P336498	
	EPA 365.1 Rev. 2.0	Total-P	0.026		mg P/L	P336732	
	SM 5310 B-00	Organic Carbon	3.0		mg C/L	P336704	
1950843	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.008	I	mg P/L	P336405	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P336430

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336534

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336536

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336935

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336754

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336498

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336732

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336405

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P336417

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P336693

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P336955

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-97
Batch ID: P336961

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P337094

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P337097

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P336697

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report Method Blank Results

Reference Method: SM 4500 F-C-97
Batch ID: P337104

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P336636

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: SM 5310 B-00
Batch ID: P336704

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336534

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	93.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336536

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	93.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336935

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336754

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	108		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336498

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336732

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336732

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336405

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.8		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P336693

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P336697

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	101		P	92 - 103

Reference Method: SM 4500 F-C-97
Batch ID: P337104

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.0		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P336636

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.3		P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P336704

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336534

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950836	Chloride	92.2		P	90 - 110
1950903	Chloride	104		P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336536

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950836	Sulfate	93.6		P	90 - 110
1950903	Sulfate	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336935

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950564	Ammonia-N	103	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336754

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950840	Kjeldahl Nitrogen	96.2	98.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336498

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950838	NO2NO3-N	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336732

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950873	Total-P	104	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336405

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950841	O-Phosphate-P	102	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P336697

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950904	Fluoride	98.3	99.6	P/P	91 - 105

Reference Method: SM 4500 F-C-97
Batch ID: P337104

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1953506	Fluoride	98.5	98.5	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P336636

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950311	Organic Carbon	101		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P336704

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950871	Organic Carbon	99.7	101	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P336430

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950334	Turbidity	0.810	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336534

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950903	Chloride	0.260	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336536

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950903	Sulfate	0.300	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336935

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950564	Ammonia-N	0.535	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336754

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950840	Kjeldahl Nitrogen	1.68	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336498

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950838	NO2NO3-N	0.805	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336732

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950873	Total-P	1.05	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336405

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950841	O-Phosphate-P	0.393	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P336417

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950836	Color (true)	4.84	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P336693

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950904	Total Alkalinity	0.436	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P336955

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950867	TDS	6.53	Sample	P	0 - 10

Reference Method: SM 2540 C-97
Batch ID: P336961

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1951397	TDS	5.80	Sample	P	0 - 10

Reference Method: SM 2540 D-97
Batch ID: P337097

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1951023	TSS	7.41	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P336697

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950904	Fluoride	0.999	Spike	P	0 - 3.0

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-97
Batch ID: P337104

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1953506	Fluoride	0.0	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P336636

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950311	Organic Carbon	1.13	Spike	P	0 - 20

Reference Method: SM 5310 B-00
Batch ID: P336704

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950871	Organic Carbon	0.988	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A80710
Included Lab Sample IDs: 1950841, 1950842, 1950843

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.8	99.9	P/P	90 - 110
O-Phosphate-P	99.9	100	P/P	90 - 110

Reference Method: SM 2120 B
Run ID: A80714
Included Lab Sample IDs: 1950835, 1950836, 1950837

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	103	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A80718
Included Lab Sample IDs: 1950835, 1950836, 1950837

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A80735
Included Lab Sample IDs: 1950835, 1950836, 1950837

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	97.9	98.9	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A80735

Included Lab Sample IDs: 1950835, 1950836, 1950837

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sulfate	98.0	97.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A80739

Included Lab Sample IDs: 1950838, 1950839, 1950840

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A80775

Included Lab Sample IDs: 1950838, 1950839

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	100	104	P/P	90 - 110
Organic Carbon	101	100	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A80806

Included Lab Sample IDs: 1950835, 1950836, 1950837

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	100	P/P	90 - 110
Total Alkalinity	99.0	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A80806

Included Lab Sample IDs: 1950837

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	105	108	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A80808

Included Lab Sample IDs: 1950840

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A80863

Included Lab Sample IDs: 1950838, 1950839, 1950840

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	97.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A80866

Included Lab Sample IDs: 1950838, 1950839, 1950840

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A80866
Included Lab Sample IDs: 1950838, 1950839, 1950840

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	96.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A80886
Included Lab Sample IDs: 1950838, 1950839, 1950840

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Run ID: A80962
Included Lab Sample IDs: 1950835, 1950836

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.6	99.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				0.810	
EPA 300.0 Rev. 2.1	Chloride	93.0	104 92.2		0.260	
	Sulfate	93.3	103 93.6		0.300	
EPA 350.1 Rev. 2.0	Ammonia-N	104	103 103			0.535
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	108	96.2 98.2			1.68
EPA 353.2 Rev. 2.0	NO2NO3-N	104	101 102			0.805
EPA 365.1 Rev. 2.0	Total-P	103	104 102			1.05
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.8	102 102			0.393
SM 2120 B	Color (true)				4.84	
SM 2320 B-97	Total Alkalinity	102			0.436	
SM 2540 C-97	TDS				6.53	
	TDS				5.80	
SM 2540 D-97	TSS				7.41	
SM 4500 F-C-97	Fluoride	101	98.3 99.6			0.999
	Fluoride	96.0	98.5 98.5			0.0
SM 5310 B-00	Organic Carbon	99.3	101			1.13
	Organic Carbon	96.4	99.7 101			0.988

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1950835, 1950836, 1950837
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1950835, 1950836, 1950837

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1950835, 1950836, 1950837
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1950838, 1950839, 1950840
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1950838, 1950839, 1950840
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1950838, 1950839, 1950840
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1950838, 1950839, 1950840
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1950841, 1950842, 1950843
SM 2120 B / E31780	True Color measured at 450 nm	1950835, 1950836, 1950837
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1950835, 1950836, 1950837
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1950835, 1950836, 1950837
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1950835, 1950836, 1950837
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1950835, 1950836, 1950837
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1950838, 1950839, 1950840

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	12/05/2017			12/05/2017 17:45	DeAsia Armster	1950835, 1950836, 1950837
EPA 300.0 Rev. 2.1	12/05/2017			12/06/2017	Julia Storbeck	1950835, 1950836, 1950837
EPA 350.1 Rev. 2.0	12/05/2017			12/13/2017	Sofia Elnemr	1950838, 1950839, 1950840
EPA 351.2 Rev. 2.0	12/05/2017	12/11/2017	Adrian Shelby	12/13/2017	Joseph Suarez	1950838, 1950839, 1950840
EPA 353.2 Rev. 2.0	12/05/2017			12/06/2017	Lauren Bottorf	1950838, 1950839, 1950840
EPA 365.1 Rev. 2.0	12/05/2017	12/11/2017	Sima Feizi	12/13/2017	Beverly Y. Sanders	1950838, 1950839, 1950840
EPA 365.1 Rev. 2.0 dissolved	12/05/2017			12/05/2017 15:16	Megan Treadwell	1950841
	12/05/2017			12/05/2017 15:29	Megan Treadwell	1950842, 1950843
SM 2120 B	12/05/2017			12/05/2017 14:44	Tanya Welborn	1950836
	12/05/2017			12/05/2017 14:45	Tanya Welborn	1950837
	12/05/2017			12/05/2017 14:46	Tanya Welborn	1950835
SM 2320 B-97	12/05/2017			12/09/2017	Dale L. Simmons	1950835, 1950836, 1950837
SM 2540 C-97	12/05/2017			12/08/2017	Yijie Li	1950835, 1950836, 1950837
SM 2540 D-97	12/05/2017			12/08/2017	Yijie Li	1950835, 1950836, 1950837
SM 4500 F-C-97	12/05/2017			12/09/2017	Dale L. Simmons	1950837
	12/05/2017			12/17/2017	Dale L. Simmons	1950835, 1950836
SM 5310 B-00	12/05/2017			12/08/2017	Ping Hua	1950838, 1950839, 1950840