



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

October 2017

Qualified Data

Yes

Location/STORET Station Name: Mule Creek @ SR 12 Date: 12-18-17
(mm/dd/yyyy)
STORET # G1TLHR0063 WBID: 684 Latitude: _____
(Decimal Degrees)
County: Liberty RQ: 12-2017-12-04-49 ORG ID: 21FLTLHR Longitude: _____
(Decimal Degrees)
Receiving Body of Water: Telogia Creek Field ID: G1TLHR0063-2017-06

Sampling Team Members				WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment			
<u>Jessie Tolbert</u> <u>Jade Burtchett</u>				☒	☒	☒	☒	☒	☒ Sample Container	☐ Van Dorn	☐ Pole	
									☐ Carboy	☐ Syringe		
									☐ Dipper	☐ Other		
									Equipment ID _____			
									Collection Method			
				☒ Wading			☐ Canoe/Kayak					
				☐ From Shore			☐ Electric Motor					
				☐ Other			☐ Gasoline Motor					
Water Level: Dry / Low / <u>Normal</u> / Above Normal / Flooded									Meter ID# <u>VSI 6920 118157</u>		Matrix: <u>Fresh</u> Marine / DI	
Photos: Yes / <u>No</u> Flow: No Flow / Intestinal / <u>Flowing</u> / NA									Tide: Rising / Falling / Slack / <u>NA</u>		Tide Times: High Low	
Time Zone	Time	Sample Depth (m)	Total Depth (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Secchi Disk (m)	Sp COND (µmhos/cm)	Temp. (C°)		
<u>EST</u>	<u>10:45</u>	<u>0.15</u>	<u>0.3</u>	<u>8.80</u>	<u>91.1</u>	<u>4.21</u>	<u>0.1</u>	<u>0.3s</u>	<u>21</u>	<u>17.1</u>		
CEN												

Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other _____									
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:									
Parameter Suite	Parameter Methods				Preservation	Lot#	Expiration	pH Check	
Preserved Nutrients	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☐ Ice ☐ H ₂ SO ₄			☐ <2	
Metals	☐ W-ICPMS ☐ W-ICP				☐ Ice ☐ HNO ₃			☐ <2	
Filtered Nutrients	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter				☐ Ice				
Chlorophyll	☐ CHLSUITE-W				☐ Ice				
Physical/Aggregate (Fresh)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				☐ Ice				
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY				☐ Ice				
Macroinvertebrates	☐ MI-FW-QLDC				☐ Formalin				
Periphyton	☐ ALGAE-ID				☐ Ice				
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice				
Molecular	☒ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD				☒ Ice				
Tracers	☒ W-SUC-AA-R ☒ W-UHERB-AA ☒ W-AHERB-AA				☒ Ice				
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R				☐ Ice ☐ Other				

Notes:

Secchi visible on bottom

G1TLHR0063-2017-06



WBID 684 Field ID ↑ STORET ↓
Mule Creek @
SR 12



G1TLHR0063

Signature: [Signature] Date: 12-18-17

Field Parameters Entered into LIMS: [Signature] 12/18/17 (Initials/Date)
Date Migrated to STORET: _____ (Initials/Date)
Field Parameters QC in LIMS: [Signature] 1/3/18 (Initials/Date)
Field Sheets Scanned/Saved: _____ (Initials/Date)

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page

Meter ID: YSI6920 119157

Project: SMP 2017

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	12-18-17	805	21.2	8.88	8.89	100	N/A	1.02	P	L
ICV	"	"	806	21.2	8.88	8.88	100	N/A	1.02	P	L
CCV	"	"	1407	22.1	8.73	8.77	100	N/A	1.02	P	L
CCV	"	"									

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	12-18-17	809	171031A	11/2018	1000	1000	P	L
ICV	"	"	810	2710787	9/2019	100	98	P	L
CCV	"	"	1410	2710787	9/2019	100	98	P	L
CCV	"	"	1411	171031A	11/2018	1000	995	P	L

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	12-18-17	813	171121A	6/2019	7.0	7.0	-52	P	L
Calibr.	"	"	815	171121C	12/2018	4.0	4.0	121.2	P	L
Calibr.	"	"	818	171121B	6/2019	10.0	9.9	-220	P	L
ICV	"	"	820	171121B	6/2019	10.0	10.0	-221	P	L
CCV	"	"	1413	171121C	12/2018	4.0	4.1	117	P	L
CCV	"	"								

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater