



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

October 2017

Qualified Data

Yes

Location/STORET Station Name: Sweetwater Creek @ CR 270

Date: 12-18-17
(mm/dd/yyyy)

STORET #: G2TLHR0018

WBID: 728

Latitude:

(Decimal Degrees)

County: Liberty RQ: 2017-12-04-49

ORG ID: 21FLTLHR

Longitude:

(Decimal Degrees)

Receiving Body of Water: Apalachicola

Field ID: G2TLHR0018-2017-04

Sampling Team Members

Jessie Tolbert
Jade Burdette

Sampling Equipment

☒ Sample Container ☐ Van Dorn ☐ Pole
☐ Carboy ☐ Syringe
☐ Dipper ☐ Other

Equipment ID

Collection Method

☒ Wading ☐ Canoe/Kayak
☐ From Shore ☐ Electric Motor
☐ Other ☐ Gasoline Motor

Water Level: Dry / Low / Normal / Above Normal / Flooded

Meter ID# YSI 6920 118157

Matrix: Fresh / Marine / DI

Photos: Yes / No

Flow: No Flow / Intestinal / Flowing NA

Tide: Rising / Falling / Slack NA

Tide Times: High Low

Time Zone	Time	Sample Depth (m)	Total Depth (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Secchi Disk (m)	Sp COND (µmhos/cm)	Temp. (C°)
EST	11:10	0.3	0.6	8.77	41.0	5.45	.01	0.65	28	17.1
CEN										

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☐ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☐ Ice ☐ H2SO4			☐ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☐ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
Chlorophyll	☐ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh)	☐ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ Coli ☐ F Coli ☐ Enterococci	☒ Ice			
Molecular	☒ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD	☒ Ice			
Tracers	☒ W-SUC-AA-R ☒ W-UHERB-AA ☒ W-AHERB-AA	☒ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R	☐ Ice ☐ Other			

Notes:

Secchi visible on bottom

G2TLHR0018-2017-04



WBID 728

Field ID ↑

STORET ↓

Sweetwater Creek @
CR270



G2TLHR0018

Signature: JT A

Date: 12-18-17

Field Parameters Entered into LIMS: JW 12/23/17

(Initials/Date)

Field Parameters QC in LIMS: JB 1/3/18

(Initials/Date)

Date Migrated to STORET:

(Initials/Date)

Field Sheets Scanned/Saved:

(Initials/Date)

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: YSI 6920 119157

Project: SMR 2017

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	12-18-17	805	21.2	8.88	8.89	100	N/A	1.02	P	L
ICV	"	"	806	21.2	8.88	8.88	100	N/A	1.02	P	L
CCV	"	"	1407	22.1	8.73	8.77	100	N/A	1.02	P	L
CCV	"	"									

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	12-18-17	809	171031A	11/2018	1000	1000	P	L
ICV	"	"	810	2710787	9/2019	100	98	P	L
CCV	"	"	1410	2710787	9/2019	100	98	P	L
CCV	"	"	1411	171031A	11/2018	1000	995	P	L

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	12-18-17	813	171121A	6/2019	7.0	7.0	-52	P	L
Calibr.	"	"	815	171121C	12/2018	4.0	4.0	121.2	P	L
Calibr.	"	"	818	171121B	6/2019	10.0	9.9	-220	P	L
ICV	"	"	820	171121B	6/2019	10.0	10.0	-221	P	L
CCV	"	"	1413	171121C	12/2018	4.0	4.1	117	P	L
CCV	"	"								

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater