



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

March 2017

PAGE 1 OF

Data Qualified?

YES / No

Location/STORET Station Name: Bear Creek @ SR 267

Date: 11-21-2017
(mm/dd/yyyy)

STORET # G1TLHR0052

WBID: 757

Latitude:

County: Gadsden RQ - 2017-10-30-55

ORG ID: 21FTLHR

Longitude:

Receiving Body of Water: LAKE Talquin

Field ID: G1TLHR0052-2017-09

Sampling Team Members

Jessie Tolbert
Paul Blair

Sampling Equipment

☒ Sample Container ☐ Van Dorn ☐ Pole
☐ Carboy ☒ Syringe
☐ Dipper ☐ Other

Equipment ID

Collection Method

☒ Wading ☐ Via Boat
☐ From Shore ☐ Canoe/Kayak
☐ Other (note below) ☐ Electric Motor
☐ Gasoline Motor

Water Level: Dry / Pooled / Low / Normal / High / Flooded

Matrix: Fresh / Marine / DI

Meter ID# 12 YSI 600 XLM 2.1 Photos: Yes / No

Tide Falling: Yes / No / NA

Time Zone	Time	Sample Depth (m)	Total Depth (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (ppt)	Secchi Disk (m)	Sp COND (µmhos/cm)	Temp. (C°)
EST	13:30	0.15	0.3	9.98	95.7	5.42	0.1	0.35	27	13.53
CEN										

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other

SBIO ID Numbers: SBIO-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Acid Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / ☒ W-NO2NO3 / ☒ W-TKN / ☒ W-TOC / ☒ W-S-A-TP	☒ Ice ☒ H2SO4	7163036	6/12/18	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☒ Ice			
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / ☒ W-F / ☒ W-TSS / ☒ TURBIDITY / ☒ W-Cl-IC / ☒ W-COLOR / ☒ W-SO4-IC / ☒ W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / ☐ W-F / ☐ W-TSS / ☐ TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ Coli ☐ F Coli ☐ Enterococci ☒ qPCR	☒ Ice			
Tracers	☒ W-SUC-AA-R ☒ W-UHERB-AA ☒ W-AHERB-AA	☒ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-AHERB-AA ☐ W-PYR-AA-R ☐ W-PCL-SQ-R ☐ W-UHERB-AA ☐ W-GLYPH ☐ W-SUC-AA-R	☐ Ice ☐ Other			

Notes:

Secchi visible on bottom

G1TLHR0052-2017-09

WBID 757

Field ID ↑

STORET ↓

Bear Creek @
SR 267

G1TLHR0052

Signature: JT

Date: 11-21-2017

Field Parameters Entered into LIMS: JB 12/7/17
(Initials/Date)

Field Parameters QC in LIMS: Jan 12/13/17
(Initials/Date)

Date Migrated to STORET: (Initials/Date)

Field Sheets Scanned/Saved: (Initials/Date)

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: YSI 6920 118157

RQ-

Project: SMP2017

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	11-21-2017	820	20.2	9.1	9.1	100	N/A	1.0	P/F	L/F
ICV	"	"	823	20.1	9.1	9.1	100	N/A	1.0	P/F	L/F
CCV	"	"	1456	19.3	9.2	9.2	100	N/A	1.0	P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	11-21-2017	826	171031A	11/2018	1000	1000	P/F	L/F
ICV	"	"	829	2707083	7/2019	100	100	P/F	L/F
CCV	"	"	14:59	2707083	7/2019	100	100	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Jessie Tolbert	11-21-2017	832	170822C	3/2019	7.0	7.0	-45	P/F	L/F
Calibr.	"	"	835	170926A	10/2018	4.0	4.0	128	P/F	L/F
Calibr.	"	"	838	170822D	3/2019	10.0	10.0	-213	P/F	L/F
ICV	"	"	840	170822D	3/2019	10.0	10.0	-213	P/F	L/F
CCV	"	"	1505	170926A	10/2018	4.0	4.0	128	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 168, slope from 4 to 7 173 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: