

Chemical Analysis Report

WQAP-2017-07-27-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **SMP 2016 Freshwater Kit A**
Request ID: **RQ-2017-06-26-38**
Customer: **WQAP**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
MS #3560
Tallahassee, FL 32399-2400
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 14-AUG-2017 13:57

A handwritten signature in black ink, appearing to read "Colin Wright", is placed on a light-colored rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: BLUE HOLE SPRING

Collection Date/Time: 07/27/2017 11:45

Field ID: G2TLHR0009-2017-02

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1917171	EPA 180.1 Rev. 2.0	Turbidity	1.1		NTU	P329275	
	EPA 300.0 Rev. 2.1	Chloride	3.4		mg Cl/L	P329389	
	SM 2120 B	Color (true)	2.5	U	PCU	P329266	
	SM 2320 B-97	Total Alkalinity	167	A	mg CaCO3/L	P329457	
	SM 2540 C-97	TDS	172	A	mg/L	P329910	
	SM 2540 D-97	TSS	2	UA	mg/L	P329593	
	SM 4500 F-C-97	Fluoride	0.076	I	mg F/L	P329461	
1917172	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P329493	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.19	I	mg N/L	P329409	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.44		mg N/L	P329437	
	EPA 365.1 Rev. 2.0	Total-P	0.026		mg P/L	P329414	
	SM 5310 B-00	Organic Carbon	0.7	I	mg C/L	P329792	
1917173	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.012		mg P/L	P329250	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P329275

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P329389

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P329493

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P329409

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P329437

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329414

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329250

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P329266

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P329457

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P329910

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P329593

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P329461

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P329792

Component	Result	Code	Units
Organic Carbon	0.5	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	93.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329493

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.2		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329409

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329437

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329414

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	96.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329250

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P329457

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P329461

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.1		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P329792

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.0		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917284	Chloride	96.1		P	90 - 110
1917345	Chloride	96.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329493

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917135	Ammonia-N	98.8	99.1	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329409

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917124	Kjeldahl Nitrogen	104	113	P/F	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329437

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917121	NO2NO3-N	104	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329414

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917106	Total-P	98.4	99.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329250

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917137	O-Phosphate-P	97.1	99.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P329461

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917171	Fluoride	96.2	95.6	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P329792

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917416	Organic Carbon	95.5	94.4	P/P	85 - 115

Quality Assurance Report Precision

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P329275

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917133	Turbidity	4.00	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329389

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917284	Chloride	0.577	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329493

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917135	Ammonia-N	0.316	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329409

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917124	Kjeldahl Nitrogen	8.35	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329437

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917121	NO2NO3-N	0.952	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329414

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917106	Total-P	1.21	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329250

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917137	O-Phosphate-P	2.34	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P329266

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917178	Color (true)	1.38	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P329457

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917171	Total Alkalinity	0.199	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P329910

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917171	TDS	1.75	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P329461

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917171	Fluoride	0.498	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P329792

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917416	Organic Carbon	0.565	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A77948
Included Lab Sample IDs: 1917171

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	103	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A77996
Included Lab Sample IDs: 1917173

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	96.6	97.6	P/P	90 - 110

Reference Method: SM 2120 B
Run ID: A78002
Included Lab Sample IDs: 1917171

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	104	105	P/P	85 - 115

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A78004

Included Lab Sample IDs: 1917171

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.5	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A78075

Included Lab Sample IDs: 1917172

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A78079

Included Lab Sample IDs: 1917171

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A78079

Included Lab Sample IDs: 1917171

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	96.1	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A78080

Included Lab Sample IDs: 1917172

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.5	97.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A78091

Included Lab Sample IDs: 1917172

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A78102

Included Lab Sample IDs: 1917172

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	101	105	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A78233

Included Lab Sample IDs: 1917172

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.3	97.0	P/P	90 - 110

Quality Assurance Report Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	SMP
EPA 180.1 Rev. 2.0	Turbidity					4.00
EPA 300.0 Rev. 2.1	Chloride	93.7	96.1	96.1		0.577
EPA 350.1 Rev. 2.0	Ammonia-N	99.2	98.8	99.1		
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	104	113		0.316
EPA 353.2 Rev. 2.0	NO2NO3-N	105	104	106		8.35
EPA 365.1 Rev. 2.0	Total-P	96.8	98.4	99.6		0.952
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100	97.1	99.4		1.21
SM 2120 B	Color (true)					2.34
SM 2320 B-97	Total Alkalinity	102			1.38	
SM 2540 C-97	TDS				0.199	
SM 4500 F-C-97	Fluoride	96.1	96.2	95.6	1.75	
SM 5310 B-00	Organic Carbon	96.0	95.5	94.4		0.498
						0.565

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1917171
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1917171
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1917172
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1917172
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1917172
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1917172
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1917173
SM 2120 B / E31780	True Color measured at 450 nm	1917171
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1917171
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1917171
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1917171
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1917171
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1917172

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/27/2017			07/27/2017 17:18	DeAsia Armster	1917171
EPA 300.0 Rev. 2.1	07/27/2017			07/28/2017	Julia Storbeck	1917171
EPA 350.1 Rev. 2.0	07/27/2017			08/01/2017	Sofia Elnemr	1917172
EPA 351.2 Rev. 2.0	07/27/2017	07/31/2017	Adrian Shelby	08/02/2017	Joseph Suarez	1917172
EPA 353.2 Rev. 2.0	07/27/2017			08/01/2017	Beverly Y. Sanders	1917172
EPA 365.1 Rev. 2.0	07/27/2017	07/31/2017	Adrian Shelby	08/01/2017	Lipi Saha	1917172
EPA 365.1 Rev. 2.0 dissolved	07/27/2017			07/27/2017 15:51	Ping Hua	1917173
SM 2120 B	07/27/2017			07/27/2017 17:00	Tanya Welborn	1917171
SM 2320 B-97	07/27/2017			08/01/2017	Dale L. Simmons	1917171
SM 2540 C-97	07/27/2017			08/01/2017	DeAsia Armster	1917171
SM 2540 D-97	07/27/2017			08/01/2017	DeAsia Armster	1917171
SM 4500 F-C-97	07/27/2017			08/01/2017	Dale L. Simmons	1917171
SM 5310 B-00	07/27/2017			08/04/2017	Sofia Elnemr	1917172