

Chemical Analysis Report

WQAP-2017-07-18-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SMP 2016 Marine Kit B, _dup & field blank**
Request ID: **RQ-2017-05-08-44**
Customer: **WQAP**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
MS #3560
Tallahassee, FL 32399-2400
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 03-AUG-2017 13:37



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Direct Runoff to Gulf

Collection Date/Time: 07/18/2017 13:15

Field ID: G1TLHR0013-2017-03

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1914191	EPA 180.1 Rev. 2.0	Turbidity	9.2		NTU	P328795	
	SM 2320 B-97	Total Alkalinity	164		mg CaCO3/L	P328816	
	SM 2540 D-97	TSS	19		mg/L	P329301	
	SM 4500 F-C-97	Fluoride	0.90		mg F/L	P328818	
1914192	EPA 350.1 Rev. 2.0	Ammonia-N	0.021		mg N/L	P328885	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P328893	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P328990	
	EPA 365.1 Rev. 2.0	Total-P	0.070		mg P/L	P328906	
	SM 5310 B-00	Organic Carbon	13		mg C/L	P329125	
1914193	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P328750	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P328795

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P328885

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P328893

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P328990

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P328906

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P328750

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P328816

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-97
Batch ID: P329301

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P328818

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P329125

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P328885

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P328893

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P328990

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P328906

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.9		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P328750

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P328750

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	105		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P328816

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P328818

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.4		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P329125

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.6		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P328893

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1914421	Kjeldahl Nitrogen	103	98.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P328990

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1913960	NO2NO3-N	102	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P328906

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1913969	Total-P	99.0	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P328750

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1914193	O-Phosphate-P	96.2	96.1	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P328818

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P328818

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1913968	Fluoride	95.0	96.5	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P329125

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1914352	Organic Carbon	93.7	92.1	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P328795

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1914213	Turbidity	5.75	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P328885

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1913960	Ammonia-N	0.245	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P328893

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1914421	Kjeldahl Nitrogen	3.54	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P328990

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1913960	NO2NO3-N	0.927	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P328906

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1913969	Total-P	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P328750

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P328750

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1914193	O-Phosphate-P	0.104	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P328816

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1913968	Total Alkalinity	0.240	Sample	P	0 - 4.8

Reference Method: SM 4500 F-C-97
Batch ID: P328818

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1913968	Fluoride	1.01	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P329125

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1914352	Organic Carbon	1.43	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A77804
Included Lab Sample IDs: 1914193

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.8	101	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A77821
Included Lab Sample IDs: 1914191

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	99.5	P/P	95 - 105

Reference Method: SM 2320 B-97
Run ID: A77832
Included Lab Sample IDs: 1914191

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97
Run ID: A77832
Included Lab Sample IDs: 1914191

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.5	99.5	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A77842
Included Lab Sample IDs: 1914192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.0	99.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A77856
Included Lab Sample IDs: 1914192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.3	99.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A77897
Included Lab Sample IDs: 1914192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	106	109	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A77900
Included Lab Sample IDs: 1914192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	95.5	97.5	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A77952
Included Lab Sample IDs: 1914192

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.3	97.8	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				5.75	
EPA 350.1 Rev. 2.0	Ammonia-N	104				0.245
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	103	98.2		3.54
EPA 353.2 Rev. 2.0	NO2NO3-N	104	102	101		0.927

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 365.1 Rev. 2.0	Total-P	97.9	99.0 99.0			0.0
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	105	96.2 96.1			0.104
SM 2320 B-97	Total Alkalinity	102			0.240	
SM 4500 F-C-97	Fluoride	97.4	95.0 96.5			1.01
SM 5310 B-00	Organic Carbon	98.6	93.7 92.1			1.43

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1914191
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1914192
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1914192
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1914192
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1914192
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1914193
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1914191
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1914191
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1914191
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1914192

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/18/2017			07/19/2017 16:03	DeAsia Armster	1914191
EPA 350.1 Rev. 2.0	07/18/2017			07/20/2017	Sofia Elnemr	1914192
EPA 351.2 Rev. 2.0	07/18/2017	07/21/2017	Adrian Shelby	07/24/2017	Joseph Suarez	1914192
EPA 353.2 Rev. 2.0	07/18/2017			07/20/2017	Beverly Y. Sanders	1914192
EPA 365.1 Rev. 2.0	07/18/2017	07/21/2017	Vijayalakshmi Reddy	07/24/2017	Lipi Saha	1914192
EPA 365.1 Rev. 2.0 dissolved	07/18/2017			07/19/2017 11:05	Ping Hua	1914193
SM 2320 B-97	07/18/2017			07/20/2017	Dale L. Simmons	1914191
SM 2540 D-97	07/18/2017			07/21/2017	Yijie Li	1914191
SM 4500 F-C-97	07/18/2017			07/20/2017	Dale L. Simmons	1914191
SM 5310 B-00	07/18/2017			07/25/2017	Ping Hua	1914192