

Chemical Analysis Report

WQAP-2017-06-27-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SMP 2016 Freshwater Kit A**
Request ID: **RQ-2017-05-08-45**
Customer: **WQAP**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
MS #3560
Tallahassee, FL 32399-2400
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 14-JUL-2017 11:33



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Owl Creek @ HWY 357

Collection Date/Time: 06/27/2017 11:45

Field ID: G1TLHR0010-2017-02

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1909352	EPA 180.1 Rev. 2.0	Turbidity	1.7		NTU	P327595	
	EPA 300.0 Rev. 2.1	Chloride	7.2		mg Cl/L	P327935	
	SM 2120 B	Color (true)	310	A	PCU	P327649	
	SM 2320 B-97	Total Alkalinity	91		mg CaCO3/L	P327907	
	SM 2540 C-97	TDS	180	A	mg/L	P328010	
	SM 2540 D-97	TSS	4	IA	mg/L	P328280	
	SM 4500 F-C-97	Fluoride	0.13		mg F/L	P327912	
1909353	EPA 350.1 Rev. 2.0	Ammonia-N	0.041		mg N/L	P327681	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P327757	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.049		mg N/L	P327700	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P327792	
	SM 5310 B-00	Organic Carbon	31		mg C/L	P327761	
1909354	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.079		mg P/L	P327641	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P327595

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P327935

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P327681

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P327757

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P327700

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P327792

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P327641

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P327649

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P327907

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P328010

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P328280

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P327912

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P327761

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P327935

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P327935

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P327681

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	109		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P327757

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P327700

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P327792

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	93.1		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P327641

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.9		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P327907

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P327912

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.3		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P327761

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P327935

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909244	Chloride	102		P	90 - 110
1909265	Chloride	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P327681

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909359	Ammonia-N	99.1	96.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P327757

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909371	Kjeldahl Nitrogen	96.3	90.7	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P327700

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909272	NO2NO3-N	100	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P327792

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909272	Total-P	102	98.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P327641

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909373	O-Phosphate-P	93.4	95.1	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P327912

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909653	Fluoride	97.8	101	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P327761

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1909072	Organic Carbon	98.6	100	P/P	85 - 115

Quality Assurance Report Precision

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P327595

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909347	Turbidity	9.32	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P327935

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909244	Chloride	3.56	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P327681

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909359	Ammonia-N	1.90	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P327757

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909371	Kjeldahl Nitrogen	5.01	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P327700

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909272	NO2NO3-N	1.48	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P327792

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909272	Total-P	3.47	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P327641

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909373	O-Phosphate-P	1.25	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P327649

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909352	Color (true)	1.30	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P327907

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909653	Total Alkalinity	0.0527	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P328010

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909352	TDS	2.79	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P327912

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909653	Fluoride	2.39	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P327761

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1909072	Organic Carbon	1.12	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A77382
Included Lab Sample IDs: 1909352

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	99.0	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A77396
Included Lab Sample IDs: 1909354

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.3	P/P	90 - 110

Reference Method: SM 2120 B
Run ID: A77400
Included Lab Sample IDs: 1909352

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	103	99.9	P/P	85 - 115

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A77410

Included Lab Sample IDs: 1909353

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	95.5	96.1	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A77418

Included Lab Sample IDs: 1909353

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	99.5	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A77437

Included Lab Sample IDs: 1909353

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	94.9	95.0	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A77490

Included Lab Sample IDs: 1909353

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	109	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A77508

Included Lab Sample IDs: 1909352

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A77508

Included Lab Sample IDs: 1909352

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	97.2	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A77522

Included Lab Sample IDs: 1909352

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	95.1	95.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A77542

Included Lab Sample IDs: 1909353

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	104	106	P/P	90 - 110

Quality Assurance Report Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision		MS
					LCS	SMP	
EPA 180.1 Rev. 2.0	Turbidity					9.32	
EPA 300.0 Rev. 2.1	Chloride	99.2	102	103		3.56	
EPA 350.1 Rev. 2.0	Ammonia-N	109	99.1	96.5			1.90
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	96.3	90.7			5.01
EPA 353.2 Rev. 2.0	NO2NO3-N	104	100	102			1.48
EPA 365.1 Rev. 2.0	Total-P	93.1	102	98.0			3.47
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.9	93.4	95.1			1.25
SM 2120 B	Color (true)					1.30	
SM 2320 B-97	Total Alkalinity	103				0.0527	
SM 2540 C-97	TDS					2.79	
SM 4500 F-C-97	Fluoride	96.3	97.8	101			2.39
SM 5310 B-00	Organic Carbon	97.8	98.6	100			1.12

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1909352
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1909352
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1909353
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1909353
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1909353
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1909353
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1909354
SM 2120 B / E31780	True Color measured at 450 nm	1909352
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1909352
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1909352
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1909352
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1909352
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1909353

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/27/2017			06/27/2017 17:22	DeAsia Armster	1909352
EPA 300.0 Rev. 2.1	06/27/2017			06/30/2017	Julia Storbeck	1909352
EPA 350.1 Rev. 2.0	06/27/2017			06/28/2017	Sofia Elnemr	1909353
EPA 351.2 Rev. 2.0	06/27/2017	06/30/2017	Adrian Shelby	07/03/2017	Joseph Suarez	1909353
EPA 353.2 Rev. 2.0	06/27/2017			06/29/2017	Beverly Y. Sanders	1909353
EPA 365.1 Rev. 2.0	06/27/2017	06/30/2017	Vijayalakshmi Reddy	07/06/2017	Lipi Saha	1909353
EPA 365.1 Rev. 2.0 dissolved	06/27/2017			06/28/2017 12:12	Lipi Saha	1909354
SM 2120 B	06/27/2017			06/28/2017 16:30	Tanya Welborn	1909352
SM 2320 B-97	06/27/2017			07/01/2017	Dale L. Simmons	1909352
SM 2540 C-97	06/27/2017			06/30/2017	Yijie Li	1909352
SM 2540 D-97	06/27/2017			06/30/2017	Yijie Li	1909352
SM 4500 F-C-97	06/27/2017			07/01/2017	Dale L. Simmons	1909352
SM 5310 B-00	06/27/2017			06/29/2017	Sofia Elnemr	1909353