

Chemical Analysis Report

WQAP-2017-07-25-07

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **pesticide template kit: C**
Request ID: **RQ-2017-06-19-34**
Customer: **WQAP**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
MS #3560
Tallahassee, FL 32399-2400
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 16-AUG-2017 11:44

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light-colored rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Mearson Spring

Collection Date/Time: 07/25/2017 12:15

Field ID: G1TLHR0025-2017-02

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1916219	EPA 180.1 Rev. 2.0	Turbidity	0.20		NTU	P329160	
	EPA 300.0 Rev. 2.1	Chloride	7.5		mg Cl/L	P329192	
		Sulfate	19		mg SO4/L	P329194	
	SM 2120 B	Color (true)	2.5	U	PCU	P329077	
	SM 2320 B-97	Total Alkalinity	172	A	mg CaCO3/L	P329382	
	SM 2540 C-97	TDS	222		mg/L	P329473	
	SM 2540 D-97	TSS	2	U	mg/L	P329781	
	SM 4500 F-C-97	Fluoride	0.12		mg F/L	P329384	
1916220	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P329231	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.13	I	mg N/L	P329173	
	EPA 353.2 Rev. 2.0	NO2NO3-N	4.0		mg N/L	P329323	
	EPA 365.1 Rev. 2.0	Total-P	0.024		mg P/L	P329225	
	SM 5310 B-00	Organic Carbon	0.66	I	mg C/L	P329396	
1916221	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.020		mg P/L	P329158	
1916226	EPA 200.7 Rev. 4.4	Calcium	69.1		mg/L	P329237	
		Iron	30	U	ug/L	P329237	
		Magnesium	11.7		mg/L	P329237	
		Potassium	1.8		mg/L	P329237	
		Sodium	3.4		mg/L	P329237	
	EPA 200.8 Rev. 5.4	Zinc	5.0	U	ug/L	P329237	
		Arsenic	0.58		ug/L	P329237	
		Cadmium	0.038	I	ug/L	P329237	
		Chromium	0.62	I	ug/L	P329237	
		Copper	0.20	U	ug/L	P329237	
		Lead	0.050	U	ug/L	P329237	
		Nickel	0.20	U	ug/L	P329237	
		Silver	0.010	U	ug/L	P329237	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Mearson Spring

Collection Date/Time: 07/25/2017 12:15

Field ID: G1TLHR0025-2017-02

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1916223	EPA 8321B	Glyphosate**	0.50	U	ug/L	P329109	
		Pyraclostrobin**	0.0020	U	ug/L	P329108	MS, RPD
		Sucralose**	0.010	U	ug/L	P329067	
	USGS O-2060-01	2,4-D	0.0020	U	ug/L	P329107	MS
		Diuron	8.0E-04	U	ug/L	P329107	
		Fenuron	0.0080	U	ug/L	P329107	
		Imidacloprid	8.0E-04	U	ug/L	P329107	
		Bentazon	8.0E-04	U	ug/L	P329107	MS
		Linuron	0.0040	U	ug/L	P329107	

Field ID: G1TLHR0025-2017-02

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1916223	USGS O-2060-01	Acetaminophen**	0.0080	U	ug/L	P329107	
		MCP	0.0020	U	ug/L	P329107	
		Carbamazepine**	4.0E-04	U	ug/L	P329107	
		Triclopyr**	0.0040	U	ug/L	P329107	
		Primidone**	0.0040	U	ug/L	P329107	
		Fluridone**	4.0E-04	U	ug/L	P329107	MS, RPD
1916224	EPA 8270D	Aldrin	4.5	U	ng/L	P329202	
		Alpha-BHC	2.2	U	ng/L	P329202	
		Beta-BHC	4.5	U	ng/L	P329202	
		Delta-BHC	4.5	U	ng/L	P329202	
		Gamma-BHC	4.5	U	ng/L	P329202	
		Carbophenothion	4.5	U	ng/L	P329202	
		Chlorothalonil	2.2	UJ	ng/L	P329202	RPD
		Cypermethrin	22	U	ng/L	P329202	
		DDD-p,p'	2.2	U	ng/L	P329202	
		DDE-p,p'	3.4	U	ng/L	P329202	
		DDT-p,p'	3.4	U	ng/L	P329202	
		Dieldrin	4.5	U	ng/L	P329202	
		Endosulfan I	4.5	U	ng/L	P329202	
		Endosulfan II	4.5	U	ng/L	P329202	
		Endosulfan Sulfate	2.2	U	ng/L	P329202	
		Endrin	2.2	U	ng/L	P329202	RPD
		Endrin Aldehyde	2.2	U	ng/L	P329202	
		Heptachlor	1.7	U	ng/L	P329202	
		Heptachlor Epoxide	2.2	U	ng/L	P329202	
		Methoxychlor	4.5	U	ng/L	P329202	
		Mirex	2.2	U	ng/L	P329202	
		Permethrin	11	U	ng/L	P329202	
		Trifluralin	1.1	U	ng/L	P329202	
		Alpha-Chlordane	1.1	U	ng/L	P329202	
		Gamma-Chlordane	1.1	U	ng/L	P329202	
		Endrin Ketone	2.2	U	ng/L	P329202	
		Dicofol	34	U	ng/L	P329202	
		Chlordane**	2.8	U	ng/L	P329202	
		Toxaphene**	14	U	ng/L	P329202	
1916225	EPA 8270D	Alachlor	0.26	U	ng/L	P329167	
		Ametryn	0.34	U	ng/L	P329167	
		Atrazine	0.39	I	ng/L	P329167	
		Atrazine Desethyl	1.6	UJ	ng/L	P329167	RPD
		Azinphos Methyl	2.3	U	ng/L	P329167	
		Bromacil	0.52	U	ng/L	P329167	
		Butylate	0.41	UJ	ng/L	P329167	RPD
		Chlorpyrifos Ethyl	0.26	U	ng/L	P329167	
		Chlorpyrifos Methyl	0.10	UJ	ng/L	P329167	CCV
		Demeton	2.1	U	ng/L	P329167	

Field ID: G1TLHR0025-2017-02

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1916225	EPA 8270D	Diazinon	0.13	U	ng/L	P329167	
		Disulfoton	0.34	U	ng/L	P329167	
		EPTC	1.3	UJ	ng/L	P329167	RPD
		Ethion	0.16	U	ng/L	P329167	
		Ethoprop	0.10	U	ng/L	P329167	
		Fenamiphos	0.26	U	ng/L	P329167	
		Fipronil	0.26	U	ng/L	P329167	
		Fipronil Sulfide	0.16	U	ng/L	P329167	
		Fipronil Sulfone	0.16	U	ng/L	P329167	
		Hexazinone	0.52	U	ng/L	P329167	
		Malathion	0.36	U	ng/L	P329167	
		Metribuzin	0.21	UJ	ng/L	P329167	LCS, MS, RPD
		Mevinphos	0.52	U	ng/L	P329167	
		Molinate	0.31	U	ng/L	P329167	
		Norflurazon	0.52	U	ng/L	P329167	
		Pendimethalin	1.6	U	ng/L	P329167	
		Phorate	0.26	U	ng/L	P329167	
		Prometon	0.93	U	ng/L	P329167	
		Prometryn	0.21	U	ng/L	P329167	
		Simazine	0.52	U	ng/L	P329167	
		Terbufos	0.10	U	ng/L	P329167	
		Terbuthylazine	0.44	U	ng/L	P329167	
		Fonofos	0.21	U	ng/L	P329167	
		Metalaxyl	0.31	I	ng/L	P329167	
		Metolachlor	1.5		ng/L	P329167	
		Acetochlor	0.26	U	ng/L	P329167	
		Cyanazine	0.52	UJ	ng/L	P329167	RPD
		Parathion Ethyl	0.26	U	ng/L	P329167	
		Parathion Methyl	0.26	U	ng/L	P329167	

Ref. Method and Comment:

EPA 8321B: MS accuracy for sucralose could not be assessed due to high concentration of parameter in the spiked sample.

EPA 8270D: Refer to the narrative for an explanation of QC Codes.

EPA 8270D: MS accuracy for Atrazine and Ametryn not assessed due to high content of these parameters in the sample spiked. Refer to the narrative for an explanation of QC Codes.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P329160

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P329237

Component	Result	Code	Units
Calcium	0.075	U	mg/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P329237

Component	Result	Code	Units
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P329237

Component	Result	Code	Units
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.20	U	ug/L
Lead	0.050	U	ug/L
Nickel	0.20	U	ug/L
Silver	0.010	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329192

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329194

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329231

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329173

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329323

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329225

Component	Result	Code	Units
-----------	--------	------	-------

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329225

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329158

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270D
Batch ID: P329167

Component	Result	Code	Units
Acetochlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Ametryn	0.32	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Azinphos Methyl	2.2	U	ng/L
Bromacil	0.50	U	ng/L
Butylate	0.40	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Methyl	0.10	U	ng/L
Cyanazine	0.50	U	ng/L
Demeton	2.0	U	ng/L
Diazinon	0.12	U	ng/L
Disulfoton	0.32	U	ng/L
EPTC	1.2	U	ng/L
Ethion	0.15	U	ng/L
Ethoprop	0.10	U	ng/L
Fenamiphos	0.25	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fonofos	0.20	U	ng/L
Hexazinone	0.50	U	ng/L
Malathion	0.35	U	ng/L
Metaxyl	0.20	U	ng/L
Metolachlor	0.25	U	ng/L
Metribuzin	0.20	U	ng/L
Mevinphos	0.50	U	ng/L
Molinate	0.30	U	ng/L
Norflurazon	0.50	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Methyl	0.25	U	ng/L
Pendimethalin	1.5	U	ng/L
Phorate	0.25	U	ng/L
Prometon	0.90	U	ng/L
Prometryn	0.20	U	ng/L
Simazine	0.50	U	ng/L
Terbufos	0.10	U	ng/L
Terbutylazine	0.42	U	ng/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8270D
Batch ID: P329202

Component	Result	Code	Units
Aldrin	4.0	U	ng/L
Alpha-BHC	2.0	U	ng/L
Alpha-Chlordane	1.0	U	ng/L
Beta-BHC	4.0	U	ng/L
Carbophenothion	4.0	U	ng/L
Chlorothalonil	2.0	U	ng/L
Cypermethrin	20	U	ng/L
DDD-p,p'	2.0	U	ng/L
DDE-p,p'	3.0	U	ng/L
DDT-p,p'	3.0	U	ng/L
Delta-BHC	4.0	U	ng/L
Dicofol	30	U	ng/L
Dieldrin	4.0	U	ng/L
Endosulfan I	4.0	U	ng/L
Endosulfan II	4.0	U	ng/L
Endosulfan Sulfate	2.0	U	ng/L
Endrin	2.0	U	ng/L
Endrin Aldehyde	2.0	U	ng/L
Endrin Ketone	2.0	U	ng/L
Gamma-BHC	4.0	U	ng/L
Gamma-Chlordane	1.0	U	ng/L
Heptachlor	1.5	U	ng/L
Heptachlor Epoxide	2.0	U	ng/L
Methoxychlor	4.0	U	ng/L
Mirex	2.0	U	ng/L
Permethrin	10	U	ng/L
Trifluralin	1.0	U	ng/L

Reference Method: EPA 8276
Batch ID: P329202

Component	Result	Code	Units
Chlordane	2.5	U	ng/L
Toxaphene	12	U	ng/L

Reference Method: EPA 8321B
Batch ID: P329067

Component	Result	Code	Units
Sucralose	0.010	U	ug/L

Reference Method: EPA 8321B
Batch ID: P329108

Component	Result	Code	Units
Pyraclostrobin	0.0020	U	ug/L

Reference Method: EPA 8321B
Batch ID: P329109

Component	Result	Code	Units
Glyphosate	0.50	U	ug/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2120 B
Batch ID: P329077

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P329382

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P329473

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P329781

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P329384

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P329396

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: USGS O-2060-01
Batch ID: P329107

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Bentazon	8.0E-04	U	ug/L
Carbamazepine	4.0E-04	U	ug/L
Diuron	8.0E-04	U	ug/L
Fenuron	0.0080	U	ug/L
Fluridone	4.0E-04	U	ug/L
Imidacloprid	8.0E-04	U	ug/L
Linuron	0.0040	U	ug/L
MCPP	0.0020	U	ug/L
Primidone	0.0040	U	ug/L
Triclopyr	0.0040	U	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P329237

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	110		P	85 - 115
Iron	109		P	85 - 115
Magnesium	110		P	85 - 115
Potassium	104		P	85 - 115
Sodium	108		P	85 - 115
Zinc	105		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4

Batch ID: P329237

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Arsenic	106		P	85 - 115
Cadmium	104		P	85 - 115
Chromium	101		P	85 - 115
Copper	99.6		P	85 - 115
Lead	97.6		P	85 - 115
Nickel	103		P	85 - 115
Silver	101		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P329192

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P329194

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	94.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P329231

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P329173

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P329323

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P329225

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P329158

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.4		P	90 - 110

Reference Method: EPA 8270D
 Batch ID: P329167

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetochlor	89.6	89.0	P/P	78 - 118
Alachlor	88.5	86.0	P/P	77 - 119
Ametryn	92.9	88.2	P/P	61 - 136
Atrazine	92.1	87.3	P/P	73 - 120
Atrazine Desethyl	93.2	64.6	P/P	16 - 122
Azinphos Methyl	105	83.9	P/P	69 - 186
Bromacil	87.9	82.2	P/P	40 - 125
Butylate	41.7	65.0	P/P	32 - 87.0
Chlorpyrifos Ethyl	83.3	70.1	P/P	53 - 110
Chlorpyrifos Methyl	84.9	66.5	P/P	31 - 119
Cyanazine	166	63.3	P/P	36 - 224
Demeton	77.2	75.5	P/P	39 - 122
Diazinon	88.1	82.8	P/P	75 - 119
Disulfoton	87.3	79.4	P/P	42 - 139
EPTC	41.3	62.7	P/P	33 - 96.0
Ethion	81.9	74.8	P/P	63 - 127
Ethoprop	76.7	67.8	P/P	65 - 107
Fenamiphos	85.3	78.9	P/P	64 - 137
Fipronil	88.3	88.4	P/P	60 - 126
Fipronil Sulfide	81.3	74.0	P/P	58 - 124
Fipronil Sulfone	74.0	68.7	P/P	57 - 126
Fonofos	82.1	74.1	P/P	65 - 111
Hexazinone	86.7	86.4	P/P	46 - 151
Malathion	82.7	73.3	P/P	68 - 132
Metaxyl	93.9	80.4	P/P	51 - 132
Metolachlor	81.1	87.7	P/P	78 - 119
Metribuzin	155	94.0	F/P	64 - 138
Mevinphos	69.4	60.9	P/P	34 - 128
Molinate	52.5	64.9	P/P	44 - 101
Norflurazon	86.1	87.0	P/P	63 - 129
Parathion Ethyl	85.1	83.0	P/P	80 - 109
Parathion Methyl	97.4	82.5	P/P	74 - 125
Pendimethalin	80.9	73.9	P/P	48 - 134
Phorate	86.8	70.5	P/P	52 - 119
Prometon	84.0	92.0	P/P	69 - 124
Prometryn	93.4	94.6	P/P	66 - 124
Simazine	89.9	88.3	P/P	61 - 134
Terbufos	82.7	73.8	P/P	58 - 115
Terbutylazine	81.4	91.4	P/P	77 - 113

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270D
Batch ID: P329202

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldrin	58.2	51.0	P/P	26 - 88.0
Alpha-BHC	85.6	75.0	P/P	63 - 114
Alpha-Chlordane	76.9	66.5	P/P	56 - 117
Beta-BHC	100	81.8	P/P	57 - 138
Carbophenothion	106	105	P/P	36 - 148
Chlorothalonil	95.0	33.2	P/P	26 - 173
Cypermethrin	120	96.7	P/P	42 - 159
DDD-p,p'	89.2	76.7	P/P	62 - 126
DDE-p,p'	76.8	67.1	P/P	53 - 115
DDT-p,p'	88.4	74.2	P/P	54 - 117
Delta-BHC	106	87.3	P/P	68 - 158
Dicofol	29.2	28.1	P/P	25 - 114
Dieldrin	85.4	71.4	P/P	59 - 126
Endosulfan I	86.0	73.4	P/P	62 - 138
Endosulfan II	95.7	74.5	P/P	61 - 150
Endosulfan Sulfate	91.3	75.4	P/P	63 - 132
Endrin	86.6	72.5	P/P	49 - 148
Endrin Aldehyde	77.7	66.1	P/P	50 - 144
Endrin Ketone	87.8	75.2	P/P	66 - 122
Gamma-BHC	98.0	82.8	P/P	60 - 137
Gamma-Chlordane	77.1	66.0	P/P	54 - 113
Heptachlor	66.8	59.2	P/P	43 - 104
Heptachlor Epoxide	79.9	68.1	P/P	64 - 118
Methoxychlor	105	91.0	P/P	63 - 129
Mirex	72.1	60.7	P/P	39 - 93.0
Permethrin	86.5	71.0	P/P	51 - 123
Trifluralin	101	83.8	P/P	35 - 187

Reference Method: EPA 8276
Batch ID: P329202

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlordane	84.0	95.9	P/P	48 - 121
Toxaphene	95.8	115	P/P	43 - 135

Reference Method: EPA 8321B
Batch ID: P329067

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sucralose	61.6		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P329108

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Pyraclostrobin	42.5		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P329109

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Glyphosate	109		P	40 - 150

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-97
 Batch ID: P329382

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
 Batch ID: P329384

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	93.5		P	92 - 103

Reference Method: SM 5310 B-00
 Batch ID: P329396

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.6		P	85 - 115

Reference Method: USGS O-2060-01
 Batch ID: P329107

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	105		P	40 - 150
Acetaminophen	91.9		P	30 - 150
Bentazon	107		P	40 - 150
Carbamazepine	96.7		P	40 - 150
Diuron	116		P	40 - 150
Fenuron	112		P	40 - 150
Fluridone	94.1		P	40 - 150
Imidacloprid	102		P	40 - 160
Linuron	97.0		P	40 - 150
MCP	94.4		P	40 - 150
Primidone	116		P	40 - 150
Triclopyr	94.4		P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
 Batch ID: P329237

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916226	Calcium	104		P	80 - 120
1916226	Iron	111		P	80 - 120
1916226	Magnesium	107		P	80 - 120
1916226	Potassium	104		P	80 - 120
1916226	Sodium	108		P	80 - 120
1916226	Zinc	103		P	80 - 120
1916433	Calcium	106		P	80 - 120
1916433	Iron	110	112	P/P	80 - 120
1916433	Magnesium	105		P	80 - 120
1916433	Potassium	104	106	P/P	80 - 120
1916433	Sodium	107	109	P/P	80 - 120
1916433	Zinc	106	106	P/P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P329237

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916226	Arsenic	109		P	80 - 120
1916226	Cadmium	105		P	80 - 120
1916226	Chromium	100		P	80 - 120
1916226	Copper	96.8		P	80 - 120
1916226	Lead	98.0		P	80 - 120
1916226	Nickel	101		P	80 - 120
1916226	Silver	102		P	80 - 120
1916433	Arsenic	106	106	P/P	80 - 120
1916433	Cadmium	104	103	P/P	80 - 120
1916433	Chromium	98.5	97.8	P/P	80 - 120
1916433	Copper	96.5	96.6	P/P	80 - 120
1916433	Lead	96.5	96.1	P/P	80 - 120
1916433	Nickel	99.7	99.7	P/P	80 - 120
1916433	Silver	100	100	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329192

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916191	Chloride	95.8		P	90 - 110
1916193	Chloride	94.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329194

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916191	Sulfate	93.5		P	90 - 110
1916193	Sulfate	92.6		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329231

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916196	Ammonia-N	105	104	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329173

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916075	Kjeldahl Nitrogen	104	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329323

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916235	NO2NO3-N	91.1	90.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329225

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916196	Total-P	101	100	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329158

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916199	O-Phosphate-P	95.2	97.2	P/P	90 - 110

Reference Method: EPA 8270D
Batch ID: P329167

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916017	Acetochlor	83.3	91.3	P/P	74 - 123
1916017	Alachlor	80.3	93.8	P/P	73 - 124
1916017	Atrazine Desethyl	60.5	81.7	P/P	12 - 103
1916017	Azinphos Methyl	86.0	106	P/P	60 - 202
1916017	Bromacil	86.2	101	P/P	42 - 133
1916017	Butylate	31.1	43.0	P/P	10 - 85.0
1916017	Chlorpyrifos Ethyl	73.6	86.6	P/P	53 - 113
1916017	Chlorpyrifos Methyl	74.9	87.0	P/P	40 - 115
1916017	Cyanazine	149	185	P/P	22 - 210
1916017	Demeton	67.9	86.7	P/P	17 - 149
1916017	Diazinon	89.5	104	P/P	72 - 124
1916017	Disulfoton	77.7	96.0	P/P	43 - 139
1916017	EPTC	30.4	47.1	P/P	10 - 86.0
1916017	Ethion	71.4	83.0	P/P	65 - 131
1916017	Ethoprop	63.0	76.0	P/P	59 - 110
1916017	Fenamiphos	80.9	85.5	P/P	57 - 138
1916017	Fipronil	95.1	101	P/P	40 - 130
1916017	Fipronil Sulfide	73.3	80.7	P/P	36 - 128
1916017	Fipronil Sulfone	71.7	88.4	P/P	16 - 145
1916017	Fonofos	80.1	98.0	P/P	60 - 112
1916017	Hexazinone	79.3	96.4	P/P	69 - 142
1916017	Malathion	74.0	80.2	P/P	68 - 132
1916017	Metalaxyl	87.3	98.1	P/P	66 - 118
1916017	Metolachlor	88.0	93.8	P/P	71 - 124
1916017	Metribuzin	249	283	F/F	61 - 140
1916017	Mevinphos	68.1	82.5	P/P	38 - 130
1916017	Molinate	48.3	58.9	P/P	23 - 99.0
1916017	Norflurazon	83.5	88.8	P/P	39 - 139
1916017	Parathion Ethyl	77.9	89.6	P/P	75 - 111
1916017	Parathion Methyl	89.8	96.7	P/P	64 - 130
1916017	Pendimethalin	75.4	87.9	P/P	26 - 173
1916017	Phorate	89.5	94.8	P/P	43 - 127
1916017	Prometon	87.5	98.6	P/P	50 - 144
1916017	Prometryn	91.7	98.1	P/P	69 - 126
1916017	Simazine	103	122	P/P	51 - 152
1916017	Terbufos	76.7	96.2	P/P	54 - 125
1916017	Terbutylazine	78.3	92.2	P/P	76 - 113

Reference Method: EPA 8270D
Batch ID: P329202

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916300	Aldrin	57.0	53.8	P/P	22 - 97.0
1916300	Alpha-BHC	81.3	92.0	P/P	43 - 132
1916300	Alpha-Chlordane	71.6	68.4	P/P	36 - 126
1916300	Beta-BHC	95.8	95.7	P/P	41 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270D
Batch ID: P329202

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916300	Carbophenothion	84.6	87.9	P/P	13 - 196
1916300	Chlorothalonil	75.2	114	P/P	10 - 266
1916300	Cypermethrin	113	127	P/P	37 - 161
1916300	DDD-p,p'	82.7	84.8	P/P	44 - 131
1916300	DDE-p,p'	70.2	66.7	P/P	32 - 127
1916300	DDT-p,p'	81.0	79.1	P/P	45 - 116
1916300	Delta-BHC	112	124	P/P	50 - 201
1916300	Dicofol	18.9	22.9	P/P	10 - 100
1916300	Dieldrin	75.9	77.1	P/P	39 - 142
1916300	Endosulfan I	75.1	80.6	P/P	51 - 139
1916300	Endosulfan II	81.8	93.7	P/P	49 - 155
1916300	Endosulfan Sulfate	85.5	86.1	P/P	54 - 136
1916300	Endrin	43.1	69.6	P/P	38 - 141
1916300	Endrin Aldehyde	67.4	69.6	P/P	14 - 152
1916300	Endrin Ketone	87.1	87.9	P/P	60 - 128
1916300	Gamma-BHC	84.6	90.1	P/P	48 - 157
1916300	Gamma-Chlordane	71.8	68.6	P/P	35 - 123
1916300	Heptachlor	61.3	56.8	P/P	25 - 112
1916300	Heptachlor Epoxide	73.1	72.7	P/P	48 - 125
1916300	Methoxychlor	97.1	94.9	P/P	52 - 127
1916300	Mirex	68.3	66.2	P/P	25 - 108
1916300	Permethrin	82.5	85.1	P/P	46 - 133
1916300	Trifluralin	109	108	P/P	29 - 254

Reference Method: EPA 8276
Batch ID: P329202

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916016	Chlordane	90.9	93.8	P/P	44 - 134
1916016	Toxaphene	105	117	P/P	36 - 147

Reference Method: EPA 8321B
Batch ID: P329108

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1917168	Pyraclostrobin	43.9	27.8	P/F	40 - 150

Reference Method: EPA 8321B
Batch ID: P329109

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916003	Glyphosate	110	118	P/P	40 - 150

Reference Method: SM 4500 F-C-97
Batch ID: P329384

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916219	Fluoride	95.5	96.1	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P329396

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916196	Organic Carbon	93.0	93.8	P/P	85 - 115

Reference Method: USGS O-2060-01
 Batch ID: P329107

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1916003	2,4-D	150	148	F/P	40 - 150
1916003	Acetaminophen	95.9	124	P/P	30 - 150
1916003	Bentazon	38.5	32.7	F/F	40 - 150
1916003	Carbamazepine	104	97.1	P/P	40 - 150
1916003	Diuron	68.4	74.7	P/P	40 - 150
1916003	Fenuron	69.8	72.5	P/P	40 - 150
1916003	Fluridone	37.7	60.2	F/P	40 - 150
1916003	Imidacloprid	107	135	P/P	40 - 160
1916003	Linuron	70.9	81.7	P/P	40 - 150
1916003	MCCP	123	127	P/P	40 - 150
1916003	Primidone	103	112	P/P	40 - 150
1916003	Triclopyr	108	128	P/P	40 - 150

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P329160

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916191	Turbidity	2.94	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
 Batch ID: P329237

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916433	Calcium	0.561	Spike	P	0 - 20
1916433	Iron	1.43	Spike	P	0 - 20
1916433	Magnesium	1.58	Spike	P	0 - 20
1916433	Potassium	1.54	Spike	P	0 - 20
1916433	Sodium	1.17	Spike	P	0 - 20
1916433	Zinc	0.779	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
 Batch ID: P329237

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916433	Arsenic	0.589	Spike	P	0 - 20
1916433	Cadmium	0.946	Spike	P	0 - 20
1916433	Chromium	0.588	Spike	P	0 - 20
1916433	Copper	0.0720	Spike	P	0 - 20
1916433	Lead	0.379	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P329237

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916433	Nickel	0.0296	Spike	P	0 - 20
1916433	Silver	0.109	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P329192

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916191	Chloride	3.38	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P329231

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916196	Ammonia-N	1.17	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P329173

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916075	Kjeldahl Nitrogen	2.16	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P329323

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916235	NO2NO3-N	1.10	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P329225

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916196	Total-P	0.361	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P329158

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916199	O-Phosphate-P	2.08	Spike	P	0 - 20

Reference Method: EPA 8270D
Batch ID: P329167

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916017	Acetochlor	9.16	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D
Batch ID: P329167

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916017	Alachlor	15.5	Spike	P	0 - 30
1916017	Ametryn	4.08	Spike	P	0 - 30
1916017	Atrazine	9.32	Spike	P	0 - 30
1916017	Atrazine Desethyl	17.6	Spike	P	0 - 30
1916017	Azinphos Methyl	20.7	Spike	P	0 - 30
1916017	Bromacil	16.2	Spike	P	0 - 30
1916017	Butylate	31.9	Spike	F	0 - 30
1916017	Chlorpyrifos Ethyl	16.2	Spike	P	0 - 30
1916017	Chlorpyrifos Methyl	14.9	Spike	P	0 - 30
1916017	Cyanazine	21.4	Spike	P	0 - 30
1916017	Demeton	24.3	Spike	P	0 - 30
1916017	Diazinon	14.7	Spike	P	0 - 30
1916017	Disulfoton	21.1	Spike	P	0 - 30
1916017	EPTC	43.1	Spike	F	0 - 30
1916017	Ethion	15.0	Spike	P	0 - 30
1916017	Ethoprop	18.8	Spike	P	0 - 30
1916017	Fenamiphos	5.57	Spike	P	0 - 30
1916017	Fipronil	5.92	Spike	P	0 - 30
1916017	Fipronil Sulfide	7.62	Spike	P	0 - 30
1916017	Fipronil Sulfone	12.7	Spike	P	0 - 30
1916017	Fonofos	20.1	Spike	P	0 - 30
1916017	Hexazinone	13.1	Spike	P	0 - 30
1916017	Malathion	8.08	Spike	P	0 - 30
1916017	Metaxyl	11.6	Spike	P	0 - 30
1916017	Metolachlor	6.40	Spike	P	0 - 30
1916017	Metribuzin	12.6	Spike	P	0 - 30
1916017	Mevinphos	19.2	Spike	P	0 - 30
1916017	Molinate	19.8	Spike	P	0 - 30
1916017	Norflurazon	6.22	Spike	P	0 - 30
1916017	Parathion Ethyl	14.0	Spike	P	0 - 30
1916017	Parathion Methyl	7.42	Spike	P	0 - 30
1916017	Pendimethalin	15.3	Spike	P	0 - 30
1916017	Phorate	5.80	Spike	P	0 - 30
1916017	Prometon	11.9	Spike	P	0 - 30
1916017	Prometryn	6.77	Spike	P	0 - 30
1916017	Simazine	16.9	Spike	P	0 - 30
1916017	Terbufos	22.5	Spike	P	0 - 30
1916017	Terbuthylazine	16.4	Spike	P	0 - 30
LFB	Acetochlor	0.744	LCS	P	0 - 30
LFB	Alachlor	2.87	LCS	P	0 - 30
LFB	Ametryn	5.22	LCS	P	0 - 30
LFB	Atrazine	5.28	LCS	P	0 - 30
LFB	Atrazine Desethyl	36.3	LCS	F	0 - 30
LFB	Azinphos Methyl	22.7	LCS	P	0 - 30
LFB	Bromacil	6.63	LCS	P	0 - 30
LFB	Butylate	43.7	LCS	F	0 - 30
LFB	Chlorpyrifos Ethyl	17.3	LCS	P	0 - 30
LFB	Chlorpyrifos Methyl	24.3	LCS	P	0 - 30
LFB	Cyanazine	89.8	LCS	F	0 - 30
LFB	Demeton	2.18	LCS	P	0 - 30
LFB	Diazinon	6.27	LCS	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D
Batch ID: P329167

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Disulfoton	9.52	LCS	P	0 - 30
LFB	EPTC	41.1	LCS	F	0 - 30
LFB	Ethion	9.09	LCS	P	0 - 30
LFB	Ethoprop	12.3	LCS	P	0 - 30
LFB	Fenamiphos	7.70	LCS	P	0 - 30
LFB	Fipronil	0.137	LCS	P	0 - 30
LFB	Fipronil Sulfide	9.46	LCS	P	0 - 30
LFB	Fipronil Sulfone	7.37	LCS	P	0 - 30
LFB	Fonofos	10.2	LCS	P	0 - 30
LFB	Hexazinone	0.291	LCS	P	0 - 30
LFB	Malathion	12.0	LCS	P	0 - 30
LFB	Metalaxyl	15.5	LCS	P	0 - 30
LFB	Metolachlor	7.80	LCS	P	0 - 30
LFB	Metribuzin	48.9	LCS	F	0 - 30
LFB	Mevinphos	13.1	LCS	P	0 - 30
LFB	Molinate	21.1	LCS	P	0 - 30
LFB	Norflurazon	0.983	LCS	P	0 - 30
LFB	Parathion Ethyl	2.54	LCS	P	0 - 30
LFB	Parathion Methyl	16.6	LCS	P	0 - 30
LFB	Pendimethalin	9.07	LCS	P	0 - 30
LFB	Phorate	20.7	LCS	P	0 - 30
LFB	Prometon	9.14	LCS	P	0 - 30
LFB	Prometryn	1.30	LCS	P	0 - 30
LFB	Simazine	1.79	LCS	P	0 - 30
LFB	Terbufos	11.4	LCS	P	0 - 30
LFB	Terbutylazine	11.6	LCS	P	0 - 30

Reference Method: EPA 8270D
Batch ID: P329202

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916300	Aldrin	5.80	Spike	P	0 - 30
1916300	Alpha-BHC	12.4	Spike	P	0 - 30
1916300	Alpha-Chlordane	4.56	Spike	P	0 - 30
1916300	Beta-BHC	0.0843	Spike	P	0 - 30
1916300	Carbophenothion	3.77	Spike	P	0 - 30
1916300	Chlorothalonil	41.2	Spike	F	0 - 30
1916300	Cypermethrin	11.3	Spike	P	0 - 30
1916300	DDD-p,p'	2.51	Spike	P	0 - 30
1916300	DDE-p,p'	5.09	Spike	P	0 - 30
1916300	DDT-p,p'	2.49	Spike	P	0 - 30
1916300	Delta-BHC	9.42	Spike	P	0 - 30
1916300	Dicofol	19.4	Spike	P	0 - 30
1916300	Dieldrin	1.61	Spike	P	0 - 30
1916300	Endosulfan I	7.08	Spike	P	0 - 30
1916300	Endosulfan II	13.5	Spike	P	0 - 30
1916300	Endosulfan Sulfate	0.677	Spike	P	0 - 30
1916300	Endrin	46.9	Spike	F	0 - 30
1916300	Endrin Aldehyde	3.20	Spike	P	0 - 30
1916300	Endrin Ketone	0.935	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D
Batch ID: P329202

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916300	Gamma-BHC	6.22	Spike	P	0 - 30
1916300	Gamma-Chlordane	4.52	Spike	P	0 - 30
1916300	Heptachlor	7.61	Spike	P	0 - 30
1916300	Heptachlor Epoxide	0.485	Spike	P	0 - 30
1916300	Methoxychlor	2.38	Spike	P	0 - 30
1916300	Mirex	3.09	Spike	P	0 - 30
1916300	Permethrin	3.08	Spike	P	0 - 30
1916300	Trifluralin	0.970	Spike	P	0 - 30
LFB	Aldrin	13.2	LCS	P	0 - 30
LFB	Alpha-BHC	13.2	LCS	P	0 - 30
LFB	Alpha-Chlordane	14.5	LCS	P	0 - 30
LFB	Beta-BHC	20.2	LCS	P	0 - 30
LFB	Carbophenothion	0.949	LCS	P	0 - 30
LFB	Chlorothalonil	96.4	LCS	F	0 - 30
LFB	Cypermethrin	21.7	LCS	P	0 - 30
LFB	DDD-p,p'	15.2	LCS	P	0 - 30
LFB	DDE-p,p'	13.5	LCS	P	0 - 30
LFB	DDT-p,p'	17.5	LCS	P	0 - 30
LFB	Delta-BHC	19.4	LCS	P	0 - 30
LFB	Dicofol	3.94	LCS	P	0 - 30
LFB	Dieldrin	17.8	LCS	P	0 - 30
LFB	Endosulfan I	15.8	LCS	P	0 - 30
LFB	Endosulfan II	24.9	LCS	P	0 - 30
LFB	Endosulfan Sulfate	19.1	LCS	P	0 - 30
LFB	Endrin	17.7	LCS	P	0 - 30
LFB	Endrin Aldehyde	16.2	LCS	P	0 - 30
LFB	Endrin Ketone	15.5	LCS	P	0 - 30
LFB	Gamma-BHC	16.8	LCS	P	0 - 30
LFB	Gamma-Chlordane	15.6	LCS	P	0 - 30
LFB	Heptachlor	12.1	LCS	P	0 - 30
LFB	Heptachlor Epoxide	16.0	LCS	P	0 - 30
LFB	Methoxychlor	14.2	LCS	P	0 - 30
LFB	Mirex	17.2	LCS	P	0 - 30
LFB	Permethrin	19.7	LCS	P	0 - 30
LFB	Trifluralin	18.9	LCS	P	0 - 30

Reference Method: EPA 8276
Batch ID: P329202

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916016	Chlordane	3.08	Spike	P	0 - 30
1916016	Toxaphene	10.9	Spike	P	0 - 30
LFB	Chlordane	13.2	LCS	P	0 - 30
LFB	Toxaphene	17.8	LCS	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P329067

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P329067

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916053	Sucralose	8.70	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P329108

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1917168	Pyraclostrobin	44.9	Spike	F	0 - 30

Reference Method: EPA 8321B
Batch ID: P329109

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916003	Glyphosate	6.86	Spike	P	0 - 30

Reference Method: SM 2120 B
Batch ID: P329077

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916191	Color (true)	3.50	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P329382

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916219	Total Alkalinity	0.975	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P329473

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916194	TDS	7.31	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P329384

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916219	Fluoride	0.484	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P329396

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916196	Organic Carbon	0.665	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: USGS O-2060-01
Batch ID: P329107

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1916003	2,4-D	0.772	Spike	P	0 - 30
1916003	Acetaminophen	19.0	Spike	P	0 - 30
1916003	Bentazon	7.55	Spike	P	0 - 30
1916003	Carbamazepine	6.79	Spike	P	0 - 30
1916003	Diuron	8.46	Spike	P	0 - 30
1916003	Fenuron	3.82	Spike	P	0 - 30
1916003	Fluridone	41.3	Spike	F	0 - 30
1916003	Imidacloprid	7.60	Spike	P	0 - 30
1916003	Linuron	14.2	Spike	P	0 - 30
1916003	MCP	3.20	Spike	P	0 - 30
1916003	Primidone	8.76	Spike	P	0 - 30
1916003	Triclopyr	17.3	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 1916223
Field Sample ID: G1TLHR0025-2017-02

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Acetaminophen-d4	91.9	P	30 - 160
EPA 8321B	Carbamazepine-d10	125	P	30 - 160
EPA 8321B	Diuron-d6	101	P	30 - 160
EPA 8321B	Diuron-d6	118	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	112	P	40 - 150
EPA 8321B	Primidone-d5	100	P	30 - 160
EPA 8321B	Sucralose-d6	50.4	P	40 - 160
USGS O-2060-01	2,4-D-d3	83.8	P	40 - 160
USGS O-2060-01	Acetaminophen-d4	91.9	P	30 - 160
USGS O-2060-01	Carbamazepine-d10	125	P	30 - 160
USGS O-2060-01	Diuron-d6	101	P	30 - 160
USGS O-2060-01	Diuron-d6	118	P	30 - 160
USGS O-2060-01	Primidone-d5	100	P	30 - 160

Lab Sample ID: 1916224
Field Sample ID: G1TLHR0025-2017-02

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270D	Decachlorobiphenyl	72.1	P	39 - 118
EPA 8270D	Tetrachloro-m-xylene	66.8	P	32 - 103
EPA 8276	Decachlorobiphenyl	76.1	P	31 - 100

Lab Sample ID: 1916225
Field Sample ID: G1TLHR0025-2017-02

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270D	Malathion-d7	67.7	P	53 - 135

Quality Assurance Report Surrogates

Lab Sample ID: 1916225
Field Sample ID: G1TLHR0025-2017-02

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270D	Simazine-d10	103	P	71 - 126

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A77938
Included Lab Sample IDs: 1916219

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	103	P/P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A77948
Included Lab Sample IDs: 1916219

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	99.8	P/P	90 - 110
Sulfate	99.9	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A77968
Included Lab Sample IDs: 1916221

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.8	98.3	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A77969
Included Lab Sample IDs: 1916219

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	99.5	P/P	95 - 105

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A77993
Included Lab Sample IDs: 1916220

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A78018
Included Lab Sample IDs: 1916220

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A78022

Included Lab Sample IDs: 1916220

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	97.8	97.8	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A78025

Included Lab Sample IDs: 1916223

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Glyphosate	112	111	P/P	70 - 160

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A78041

Included Lab Sample IDs: 1916226

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	103	101	P/P	90 - 110
Iron	102	102	P/P	90 - 110
Magnesium	102	102	P/P	90 - 110
Potassium	99.6	100	P/P	90 - 110
Sodium	101	101	P/P	90 - 110
Zinc	102	103	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A78042

Included Lab Sample IDs: 1916226

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Arsenic	102	103	P/P	90 - 110
Cadmium	102	102	P/P	90 - 110
Chromium	99.2	98.9	P/P	90 - 110
Copper	100	100	P/P	90 - 110
Lead	94.9	95.6	P/P	90 - 110
Nickel	99.6	100	P/P	90 - 110
Silver	96.2	95.4	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A78044

Included Lab Sample IDs: 1916219

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	99.7	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A78044

Included Lab Sample IDs: 1916219

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	95.7	96.1	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A78047

Included Lab Sample IDs: 1916220

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	98.5	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A78052

Included Lab Sample IDs: 1916220

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.8	97.8	P/P	90 - 110

Reference Method: EPA 8270D

Run ID: A78147

Included Lab Sample IDs: 1916225

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	100		P	80 - 120
Alachlor	98.4		P	80 - 120
Ametryn	97.9		P	80 - 120
Atrazine	103		P	80 - 120
Atrazine Desethyl	105		P	80 - 120
Azinphos Methyl	93.5		P	80 - 120
Bromacil	95.0		P	80 - 120
Butylate	88.8		P	80 - 120
Chlorpyrifos Ethyl	85.8		P	80 - 120
Chlorpyrifos Methyl	77.6*		F	80 - 120
Cyanazine	112		P	80 - 120
Demeton	88.4		P	80 - 120
Diazinon	102		P	80 - 120
Disulfoton	85.7		P	80 - 120
EPTC	92.4		P	80 - 120
Ethion	82.6		P	80 - 120
Ethoprop	98.2		P	80 - 120
Fenamiphos	86.5		P	80 - 120
Fipronil	88.7		P	80 - 120
Fipronil Sulfide	84.3		P	80 - 120
Fipronil Sulfone	84.5		P	80 - 120
Fonofos	94.2		P	80 - 120
Hexazinone	106		P	80 - 120
Malathion	83.7		P	80 - 120
Metalaxyl	97.1		P	80 - 120
Metolachlor	97.5		P	80 - 120
Metribuzin	113		P	80 - 120
Mevinphos	92.4		P	80 - 120
Molinate	106		P	80 - 120
Norflurazon	96.3		P	80 - 120
Parathion Ethyl	94.0		P	80 - 120
Parathion Methyl	100		P	80 - 120
Pendimethalin	88.6		P	80 - 120
Phorate	101		P	80 - 120
Prometon	94.1		P	80 - 120
Prometryn	99.8		P	80 - 120
Simazine	99.6		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270D
Run ID: A78147
Included Lab Sample IDs: 1916225

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Terbufos	95.3		P	80 - 120
Terbutylazine	101		P	80 - 120

Reference Method: EPA 8270D
Run ID: A78159
Included Lab Sample IDs: 1916224

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aldrin	96.2		P	80 - 120
Alpha-BHC	98.1		P	80 - 120
Alpha-Chlordane	95.8		P	80 - 120
Beta-BHC	111		P	80 - 120
Carbophenothion	92.3		P	80 - 120
Chlorothalonil	93.9		P	80 - 120
Cypermethrin	95.9		P	80 - 120
DDD-p,p'	99.7		P	80 - 120
DDE-p,p'	99.0		P	80 - 120
DDT-p,p'	105		P	80 - 120
Delta-BHC	103		P	80 - 120
Dicofol	108		P	80 - 120
Dieldrin	101		P	80 - 120
Endosulfan I	90.4		P	80 - 120
Endosulfan II	92.9		P	80 - 120
Endosulfan Sulfate	96.8		P	80 - 120
Endrin	97.4		P	80 - 120
Endrin Aldehyde	94.2		P	80 - 120
Endrin Ketone	98.4		P	80 - 120
Gamma-BHC	106		P	80 - 120
Gamma-Chlordane	99.8		P	80 - 120
Heptachlor	94.7		P	80 - 120
Heptachlor Epoxide	98.2		P	80 - 120
Methoxychlor	107		P	80 - 120
Mirex	101		P	80 - 120
Permethrin	95.4		P	80 - 120
Trifluralin	98.3		P	80 - 120

Reference Method: EPA 8276
Run ID: A78177
Included Lab Sample IDs: 1916224

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlordane	104		P	80 - 120
Toxaphene	112		P	80 - 120

Reference Method: EPA 8321B
Run ID: A78239
Included Lab Sample IDs: 1916223

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Pyraclostrobin	80.5	81.4	P/P	60 - 150

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
Run ID: A78247
Included Lab Sample IDs: 1916223

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	88.1	60.3	P/P	60 - 160

Reference Method: USGS O-2060-01
Run ID: A78250
Included Lab Sample IDs: 1916223

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	73.4	73.7	P/P	50 - 150
Acetaminophen	87.8	93.7	P/P	60 - 150
Bentazon	122	106	P/P	50 - 150
Carbamazepine	89.6	88.2	P/P	60 - 150
Diuron	93.4	96.9	P/P	60 - 150
Fenuron	96.3	103	P/P	60 - 150
Fluridone	104	89.3	P/P	60 - 160
Imidacloprid	83.4	95.1	P/P	60 - 150
Linuron	88.7	98.1	P/P	60 - 150
MCPP	82.7	78.0	P/P	50 - 150
Primidone	97.2	94.2	P/P	60 - 150
Triclopyr	84.2	72.9	P/P	50 - 150

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity						2.94	
EPA 200.7 Rev. 4.4	Calcium	110	104	106				0.561
	Iron	109	111	110	112			1.43
	Magnesium	110	107	105				1.58
	Potassium	104	104	104	106			1.54
	Sodium	108	108	107	109			1.17
	Zinc	105	103	106	106			0.779
EPA 200.8 Rev. 5.4	Arsenic	106	109	106	106			0.589
	Cadmium	104	105	104	103			0.946
	Chromium	101	100	98.5	97.8			0.588
	Copper	99.6	96.8	96.5	96.6			0.0720
	Lead	97.6	98.0	96.5	96.1			0.379
	Nickel	103	101	99.7	99.7			0.0296
	Silver	101	102	100	100			0.109
EPA 300.0 Rev. 2.1	Chloride	97.6	95.8	94.2			3.38	
	Sulfate	94.9	93.5	92.6				
EPA 350.1 Rev. 2.0	Ammonia-N	104	105	104				1.17
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104	104	100				2.16
EPA 353.2 Rev. 2.0	NO2NO3-N	102	91.1	90.1				1.10
EPA 365.1 Rev. 2.0	Total-P	98.2	101	100				0.361
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.4	95.2	97.2				2.08

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
						SMP	
EPA 8270D	Acetochlor	89.6	89.0	83.3	91.3	0.744	9.16
	Alachlor	88.5	86.0	80.3	93.8	2.87	15.5
	Aldrin	58.2	51.0	57.0	53.8	13.2	5.80
	Alpha-BHC	85.6	75.0	81.3	92.0	13.2	12.4
	Alpha-Chlordane	76.9	66.5	71.6	68.4	14.5	4.56
	Ametryn	92.9	88.2			5.22	4.08
	Atrazine	92.1	87.3			5.28	9.32
	Atrazine Desethyl	93.2	64.6	60.5	81.7	36.3	17.6
	Azinphos Methyl	105	83.9	86.0	106	22.7	20.7
	Beta-BHC	100	81.8	95.8	95.7	20.2	0.0843
	Bromacil	87.9	82.2	86.2	101	6.63	16.2
	Butylate	41.7	65.0	31.1	43.0	43.7	31.9
	Carbophenothion	106	105	84.6	87.9	0.949	3.77
	Chlorothalonil	95.0	33.2	75.2	114	96.4	41.2
	Chlorpyrifos Ethyl	83.3	70.1	73.6	86.6	17.3	16.2
	Chlorpyrifos Methyl	84.9	66.5	74.9	87.0	24.3	14.9
	Cyanazine	166	63.3	149	185	89.8	21.4
	Cypermethrin	120	96.7	113	127	21.7	11.3
	DDD-p,p'	89.2	76.7	82.7	84.8	15.2	2.51
	DDE-p,p'	76.8	67.1	70.2	66.7	13.5	5.09
	DDT-p,p'	88.4	74.2	81.0	79.1	17.5	2.49
	Delta-BHC	106	87.3	112	124	19.4	9.42
	Demeton	77.2	75.5	67.9	86.7	2.18	24.3
	Diazinon	88.1	82.8	89.5	104	6.27	14.7
	Dicofol	29.2	28.1	18.9	22.9	3.94	19.4
	Dieldrin	85.4	71.4	75.9	77.1	17.8	1.61
	Disulfoton	87.3	79.4	77.7	96.0	9.52	21.1
	Endosulfan I	86.0	73.4	75.1	80.6	15.8	7.08
	Endosulfan II	95.7	74.5	81.8	93.7	24.9	13.5
	Endosulfan Sulfate	91.3	75.4	85.5	86.1	19.1	0.677
	Endrin	86.6	72.5	43.1	69.6	17.7	46.9
	Endrin Aldehyde	77.7	66.1	67.4	69.6	16.2	3.20
	Endrin Ketone	87.8	75.2	87.1	87.9	15.5	0.935
	EPTC	41.3	62.7	30.4	47.1	41.1	43.1
	Ethion	81.9	74.8	71.4	83.0	9.09	15.0
	Ethoprop	76.7	67.8	63.0	76.0	12.3	18.8
	Fenamiphos	85.3	78.9	80.9	85.5	7.70	5.57
	Fipronil	88.3	88.4	95.1	101	0.137	5.92
	Fipronil Sulfide	81.3	74.0	73.3	80.7	9.46	7.62
	Fipronil Sulfone	74.0	68.7	71.7	88.4	7.37	12.7
	Fonofos	82.1	74.1	80.1	98.0	10.2	20.1
	Gamma-BHC	98.0	82.8	84.6	90.1	16.8	6.22
	Gamma-Chlordane	77.1	66.0	71.8	68.6	15.6	4.52
	Heptachlor	66.8	59.2	61.3	56.8	12.1	7.61
	Heptachlor Epoxide	79.9	68.1	73.1	72.7	16.0	0.485
	Hexazinone	86.7	86.4	79.3	96.4	0.291	13.1
	Malathion	82.7	73.3	74.0	80.2	12.0	8.08
	Metalaxyl	93.9	80.4	87.3	98.1	15.5	11.6
	Methoxychlor	105	91.0	97.1	94.9	14.2	2.38
	Metolachlor	81.1	87.7	88.0	93.8	7.80	6.40
	Metribuzin	155	94.0	249	283	48.9	12.6
	Mevinphos	69.4	60.9	68.1	82.5	13.1	19.2
	Mirex	72.1	60.7	68.3	66.2	17.2	3.09
	Molinate	52.5	64.9	48.3	58.9	21.1	19.8

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
						SMP	
EPA 8270D	Norflurazon	86.1	87.0	83.5	88.8	0.983	6.22
	Parathion Ethyl	85.1	83.0	77.9	89.6	2.54	14.0
	Parathion Methyl	97.4	82.5	89.8	96.7	16.6	7.42
	Pendimethalin	80.9	73.9	75.4	87.9	9.07	15.3
	Permethrin	86.5	71.0	82.5	85.1	19.7	3.08
	Phorate	86.8	70.5	89.5	94.8	20.7	5.80
	Prometon	84.0	92.0	87.5	98.6	9.14	11.9
	Prometryn	93.4	94.6	91.7	98.1	1.30	6.77
	Simazine	89.9	88.3	103	122	1.79	16.9
	Terbufos	82.7	73.8	76.7	96.2	11.4	22.5
	Terbutylazine	81.4	91.4	78.3	92.2	11.6	16.4
	Trifluralin	101	83.8	109	108	18.9	0.970
EPA 8276	Chlordane	84.0	95.9	90.9	93.8	13.2	3.08
	Toxaphene	95.8	115	105	117	17.8	10.9
EPA 8321B	Glyphosate	109		110	118		6.86
	Pyraclostrobin	42.5		43.9	27.8		44.9
	Sucralose	61.6					8.70
SM 2120 B	Color (true)					3.50	
SM 2320 B-97	Total Alkalinity	102				0.975	
SM 2540 C-97	TDS					7.31	
SM 4500 F-C-97	Fluoride	93.5		95.5	96.1		0.484
SM 5310 B-00	Organic Carbon	99.6		93.0	93.8		0.665
USGS O-2060-01	2,4-D	105		150	148		0.772
	Acetaminophen	91.9		95.9	124		19.0
	Bentazon	107		38.5	32.7		7.55
	Carbamazepine	96.7		104	97.1		6.79
	Diuron	116		68.4	74.7		8.46
	Fenuron	112		69.8	72.5		3.82
	Fluridone	94.1		37.7	60.2		41.3
	Imidacloprid	102		107	135		7.60
	Linuron	97.0		70.9	81.7		14.2
	MCP	94.4		123	127		3.20
	Primidone	116		103	112		8.76
	Triclopyr	94.4		108	128		17.3

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1916219
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	1916226
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	1916226
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1916219
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1916219
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1916220
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1916220
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1916220

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1916220
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1916221
EPA 8270D / E31780	Organochlorine pesticides in water matrices by GC/MS-SIM	1916224
EPA 8270D / E31780	Organonitrogen and phosphorus pesticides (ultra trace level) in water matrices by GC/MS/MS	1916225
EPA 8276 / E31780	Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization	1916224
EPA 8321B / E31780	Glyphosate in water matrices by HPLC/MS/MS	1916223
EPA 8321B / E31780	Pyraclostrobin in water matrices by HPLC/MS/MS	1916223
EPA 8321B / E31780	Water soluble organic compounds by HPLC/MS/MS	1916223
SM 2120 B / E31780	True Color measured at 450 nm	1916219
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1916219
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1916219
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1916219
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1916219
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1916220
USGS O-2060-01 / E31780	Chlorinated (phenoxy acid) herbicides in water matrices by HPLC/MS/MS	1916223
USGS O-2060-01 / E31780	Urea herbicides and other chemicals in water matrices by HPLC/MS/MS	1916223

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/25/2017			07/26/2017 15:19	DeAsia Armster	1916219
EPA 200.7 Rev. 4.4	07/25/2017	07/27/2017	Elliott D. Healy	07/28/2017	Betina Topolski	1916226
EPA 200.8 Rev. 5.4	07/25/2017	07/27/2017	Elliott D. Healy	07/28/2017	Nadine Brooks	1916226
EPA 300.0 Rev. 2.1	07/25/2017			07/26/2017	Julia Storbeck	1916219
EPA 350.1 Rev. 2.0	07/25/2017			07/26/2017	Sofia Elnemr	1916220
EPA 351.2 Rev. 2.0	07/25/2017	07/27/2017	Adrian Shelby	07/28/2017	Tanya Welborn	1916220
EPA 353.2 Rev. 2.0	07/25/2017			07/28/2017	Beverly Y. Sanders	1916220
EPA 365.1 Rev. 2.0	07/25/2017	07/27/2017	Adrian Shelby	07/28/2017	Lipi Saha	1916220
EPA 365.1 Rev. 2.0 dissolved	07/25/2017			07/26/2017 15:09	Ping Hua	1916221
EPA 8270D	07/25/2017	07/27/2017	Rasheda Ghaffari	07/31/2017	Vandana T. Nagar	1916224
	07/25/2017	07/28/2017	Fe-Val Siddell	08/03/2017	Irina Carbone	1916225
EPA 8276	07/25/2017	07/27/2017	Rasheda Ghaffari	08/01/2017	Vandana T. Nagar	1916224
EPA 8321B	07/25/2017	07/27/2017	Julie Michelle Frank	08/04/2017	Mohammad Ghaffari	1916223
	07/25/2017	07/28/2017	Julie Michelle Frank	07/28/2017	Julie Michelle Frank	1916223
	07/25/2017	07/31/2017	Fe-Val Siddell	08/06/2017	Juyoung Kim	1916223
SM 2120 B	07/25/2017			07/25/2017 17:13	Tanya Welborn	1916219
SM 2320 B-97	07/25/2017			07/27/2017	Dale L. Simmons	1916219
SM 2540 C-97	07/25/2017			07/28/2017	Yijie Li	1916219
SM 2540 D-97	07/25/2017			07/28/2017	Yijie Li	1916219
SM 4500 F-C-97	07/25/2017			07/27/2017	Dale L. Simmons	1916219
SM 5310 B-00	07/25/2017			07/28/2017	Sofia Elnemr	1916220
USGS O-2060-01	07/25/2017	07/28/2017	Hoor Shaik	08/02/2017	Juyoung Kim	1916223