

Chemical Analysis Report

WQAP-2017-09-22-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **RUN 6**
Request ID: **RQ-2017-09-18-31**
Customer: **WQAP**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Tampa, FL 33637
Attn: Judy Ashton

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 11-OCT-2017 11:21

A handwritten signature in black ink, appearing to read "Colin Wright", is placed over a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: BIG VIENNA

Collection Date/Time: 09/21/2017 12:50

Field ID: G5SW0113

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1930996	EPA 180.1 Rev. 2.0	Turbidity	3.4		NTU	P332115	
	EPA 300.0 Rev. 2.1	Chloride	21		mg Cl/L	P332217	
		Sulfate	12		mg SO4/L	P332220	
	SM 2120 B	Color (true)	93		PCU	P332106	
	SM 2320 B-97	Total Alkalinity	24		mg CaCO3/L	P332210	
	SM 2540 C-97	TDS	115		mg/L	P332591	
	SM 2540 D-97	TSS	6	I	mg/L	P332585	
	SM 4500 F-C-97	Fluoride	0.074	I	mg F/L	P332211	RPD
1931000	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P332337	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P332311	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P332605	
	EPA 365.1 Rev. 2.0	Total-P	0.046		mg P/L	P332319	
	SM 5310 B-00	Organic Carbon	14		mg C/L	P332406	
1931004	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P332143	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: VIENNA N

Collection Date/Time: 09/21/2017 12:00

Field ID: G5SW0115

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1930997	EPA 180.1 Rev. 2.0	Turbidity	3.1	A	NTU	P332116	
	EPA 300.0 Rev. 2.1	Chloride	18		mg Cl/L	P332217	
		Sulfate	11		mg SO4/L	P332220	
	SM 2120 B	Color (true)	100		PCU	P332106	
	SM 2320 B-97	Total Alkalinity	30		mg CaCO3/L	P332210	
	SM 2540 C-97	TDS	108		mg/L	P332591	
	SM 2540 D-97	TSS	4	I	mg/L	P332585	
	SM 4500 F-C-97	Fluoride	0.073	I	mg F/L	P332211	RPD
1931001	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	I	mg N/L	P332337	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.89		mg N/L	P332311	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P332605	
	EPA 365.1 Rev. 2.0	Total-P	0.033		mg P/L	P332319	
	SM 5310 B-00	Organic Carbon	13		mg C/L	P332406	
1931005	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P332143	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: HIAWATHA

Collection Date/Time: 09/21/2017 09:40

Field ID: G1SW0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1930998	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P332116	
	EPA 300.0 Rev. 2.1	Chloride	25	A	mg Cl/L	P332302	
		Sulfate	5.6	A	mg SO4/L	P332303	

Field ID: G1SW0017

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1930998	SM 2120 B	Color (true)	61	A	PCU	P332106	RPD
	SM 2320 B-97	Total Alkalinity	7.9		mg CaCO3/L	P332210	
	SM 2540 C-97	TDS	80		mg/L	P332591	
	SM 2540 D-97	TSS	3	I	mg/L	P332585	
	SM 4500 F-C-97	Fluoride	0.081	I	mg F/L	P332211	
1931002	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P332337	RPD
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.66		mg N/L	P332311	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P332605	
	EPA 365.1 Rev. 2.0	Total-P	0.037		mg P/L	P332319	
	SM 5310 B-00	Organic Carbon	11		mg C/L	P332406	
1931006	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P332143	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: GREEN

Collection Date/Time: 09/21/2017 13:38

Field ID: G5SW0107

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1930999	EPA 180.1 Rev. 2.0	Turbidity	1.6		NTU	P332116	RPD
	EPA 300.0 Rev. 2.1	Chloride	14		mg Cl/L	P332302	
		Sulfate	0.41	I	mg SO4/L	P332303	
	SM 2120 B	Color (true)	45		PCU	P332106	
	SM 2320 B-97	Total Alkalinity	20		mg CaCO3/L	P332210	
	SM 2540 C-97	TDS	61		mg/L	P332591	RPD
	SM 2540 D-97	TSS	2	I	mg/L	P332585	
	SM 4500 F-C-97	Fluoride	0.058	I	mg F/L	P332211	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P332337	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.65		mg N/L	P332311	
1931003	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P332605	RPD
	EPA 365.1 Rev. 2.0	Total-P	0.024		mg P/L	P332319	
	SM 5310 B-00	Organic Carbon	9.8		mg C/L	P332406	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P332143	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P332115

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P332116

Component	Result	Code	Units
-----------	--------	------	-------

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P332116

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332217

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332220

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332302

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332303

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332337

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332311

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332605

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332319

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332143

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P332106

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P332210

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P332591

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P332585

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P332211

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P332406

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332217

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332220

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332220

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	109		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332302

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332303

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	106		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332337

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	107		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332311

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	97.5		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332605

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332319

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332143

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	107		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P332210

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	95 - 108

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P332211

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.6		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P332406

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332217

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1930961	Chloride	97.2		P	90 - 110
1931239	Chloride	97.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332220

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1930961	Sulfate	108		P	90 - 110
1931239	Sulfate	109		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332302

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1930998	Chloride	95.6		P	90 - 110
1930999	Chloride	95.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332303

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1930998	Sulfate	108		P	90 - 110
1930999	Sulfate	105		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332337

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1930963	Ammonia-N	104	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332311

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1930963	Kjeldahl Nitrogen	95.3	92.8	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332605

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931000	NO2NO3-N	106	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332319

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931052	Total-P	97.4	99.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332143

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931006	O-Phosphate-P	100	99.0	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P332211

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1931054	Fluoride	95.8	101	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P332406

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1930963	Organic Carbon	94.3	95.4	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P332115

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930887	Turbidity	4.58	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P332116

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930997	Turbidity	0.0	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332217

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931239	Chloride	0.292	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332220

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931239	Sulfate	1.12	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332302

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930998	Chloride	1.93	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P332303

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930998	Sulfate	3.49	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P332337

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930963	Ammonia-N	2.94	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P332311

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930963	Kjeldahl Nitrogen	2.27	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P332605

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931000	NO2NO3-N	0.952	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P332319

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931052	Total-P	1.88	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P332143

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931006	O-Phosphate-P	1.30	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P332106

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930998	Color (true)	1.23	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P332210

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931054	Total Alkalinity	0.0265	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P332591

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930993	TDS	7.27	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P332211

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1931054	Fluoride	4.56	Spike	F	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P332406

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1930963	Organic Carbon	0.994	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A79108
Included Lab Sample IDs: 1930996, 1930997, 1930998, 1930999

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	103	103	P/P	85 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A79112
Included Lab Sample IDs: 1930996, 1930997, 1930998, 1930999

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105
Turbidity	99.5	100	P/P	95 - 105

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A79121

Included Lab Sample IDs: 1931004, 1931005, 1931006, 1931007

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	101	P/P	90 - 110
O-Phosphate-P	102	100	P/P	90 - 110
O-Phosphate-P	99.5	101	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A79144

Included Lab Sample IDs: 1930996, 1930997, 1930998, 1930999

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	98.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A79144

Included Lab Sample IDs: 1930996, 1930997, 1930998, 1930999

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	99.5	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A79147

Included Lab Sample IDs: 1930996, 1930997, 1930998, 1930999

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	104	103	P/P	90 - 110
Chloride	106	101	P/P	90 - 110
Sulfate	105	109	P/P	90 - 110
Sulfate	106	99.0	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A79196

Included Lab Sample IDs: 1931000, 1931001, 1931002, 1931003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	104	105	P/P	90 - 110
Ammonia-N	105	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A79212

Included Lab Sample IDs: 1931000, 1931001, 1931003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	102	P/P	90 - 110
Total-P	102	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A79215

Included Lab Sample IDs: 1931000, 1931001, 1931002, 1931003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.8	93.4	P/P	90 - 110
Organic Carbon	94.0	93.8	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A79285

Included Lab Sample IDs: 1931002

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.0	99.2	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A79298

Included Lab Sample IDs: 1931000, 1931001, 1931002, 1931003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A79310

Included Lab Sample IDs: 1931000, 1931001, 1931002, 1931003

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	92.3	96.4	P/P	90 - 110
Kjeldahl Nitrogen	96.4	93.7	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity					4.58	
	Turbidity					0.0	
EPA 300.0 Rev. 2.1	Chloride	99.9	97.5	97.2		0.292	
	Chloride	98.8	95.6	95.8		1.93	
	Sulfate	109	109	108		1.12	
	Sulfate	106	108	105		3.49	
EPA 350.1 Rev. 2.0	Ammonia-N	107	104	101			2.94
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	97.5	95.3	92.8			2.27
EPA 353.2 Rev. 2.0	NO2NO3-N	104	106	104			0.952
EPA 365.1 Rev. 2.0	Total-P	98.0	97.4	99.7			1.88
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	107	100	99.0			1.30
SM 2120 B	Color (true)					1.23	
SM 2320 B-97	Total Alkalinity	101				0.0265	
SM 2540 C-97	TDS					7.27	
SM 4500 F-C-97	Fluoride	96.6	95.8	101			4.56
SM 5310 B-00	Organic Carbon	94.4	94.3	95.4			0.994

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1930996, 1930997, 1930998, 1930999
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1930996, 1930997, 1930998, 1930999

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1930996, 1930997, 1930998, 1930999
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1931000, 1931001, 1931002, 1931003
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1931000, 1931001, 1931002, 1931003
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1931000, 1931001, 1931002, 1931003
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1931000, 1931001, 1931002, 1931003
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1931004, 1931005, 1931006, 1931007
SM 2120 B / E31780	True Color measured at 450 nm	1930996, 1930997, 1930998, 1930999
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1930996, 1930997, 1930998, 1930999
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1930996, 1930997, 1930998, 1930999
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1930996, 1930997, 1930998, 1930999
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1930996, 1930997, 1930998, 1930999
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1931000, 1931001, 1931002, 1931003

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/22/2017			09/22/2017 17:40	Tanya Welborn	1930996, 1930997, 1930998, 1930999
EPA 300.0 Rev. 2.1	09/22/2017			09/26/2017	Julia Storbeck	1930996, 1930997, 1930998, 1930999
EPA 350.1 Rev. 2.0	09/22/2017			09/26/2017	Sofia Elnemr	1931000, 1931001, 1931002, 1931003
EPA 351.2 Rev. 2.0	09/22/2017	09/27/2017	Adrian Shelby	09/29/2017	Joseph Suarez	1931000, 1931001, 1931002, 1931003
EPA 353.2 Rev. 2.0	09/22/2017			10/03/2017	Beverly Y. Sanders	1931000, 1931001, 1931002, 1931003
EPA 365.1 Rev. 2.0	09/22/2017	09/27/2017	Sima Feizi	09/28/2017	Lipi Saha	1931000, 1931001, 1931003
	09/22/2017	09/27/2017	Sima Feizi	10/02/2017	Lipi Saha	1931002
EPA 365.1 Rev. 2.0 dissolved	09/22/2017			09/22/2017 15:08	Megan Treadwell	1931006
	09/22/2017			09/22/2017 15:36	Megan Treadwell	1931005
	09/22/2017			09/22/2017 15:41	Megan Treadwell	1931007
	09/22/2017			09/22/2017 15:48	Megan Treadwell	1931004
SM 2120 B	09/22/2017			09/22/2017 15:43	Julia Storbeck	1930996, 1930998
	09/22/2017			09/22/2017 15:44	Julia Storbeck	1930997
	09/22/2017			09/22/2017 15:45	Julia Storbeck	1930999
SM 2320 B-97	09/22/2017			09/25/2017	Dale L. Simmons	1930996, 1930997, 1930998, 1930999
SM 2540 C-97	09/22/2017			09/27/2017	Yijie Li	1930996, 1930997, 1930998, 1930999
SM 2540 D-97	09/22/2017			09/27/2017	Yijie Li	1930996, 1930997, 1930998, 1930999
SM 4500 F-C-97	09/22/2017			09/25/2017	Dale L. Simmons	1930996, 1930997, 1930998, 1930999
SM 5310 B-00	09/22/2017			09/27/2017	Ping Hua	1931000, 1931001, 1931002, 1931003